

CLARE MEDICAL BUILDING

101 WEST 7TH STREET, CLARE, MICHIGAN 48617

E.C. SHURLOW, D.O.
Physician and Surgeon

8-5-85

U.S. Nuclear Regulatory Commission
Regional Licensing Section
799 Roosevelt Road
Glen Ellyn, Illinois 60137


Re: License #21-17024-01

Gentlemen:

This is to notify the Commission of our desire to terminate our byproduct materials license issued to Clare Medical Diagnostics. Nuclear medicine services have been terminated. A close-out survey is enclosed.

If you have any questions, please do not hesitate to contact me.

Sincerely,



E. C. Shurlow, D.O.

enclosures

U.S. N.R.C.
LIC. FEE MGMT. BRANCH

85 AUG 14 PM 2:27

RECEIVED

RECEIVED BY LFMB	
Date	8/14/85
Log	aug 16 11
By	P. J. H.
Orig. To	P. J. H.
Action Compl.	P. J. H.

RECEIVED
AUG 8 1985
REGION III

8509100529 850820
REG3 LIC30
21-17024-01 PDR

FEE EXEMPT

AUG 8 1985

CONTROL NO. 7 9530

CLOSE-OUT SURVEY


CLARE MEDICAL DIAGNOSTICS

LICENSE NO. 21-17024-01

A close-out survey was performed in the facilities of Clare Medical Diagnostics on 7/9/85 by David Close, Consultant, Nuclear Medicine Associates. The survey consisted of a series of wipe tests and measurements with a G-M survey meter. The areas evaluated include the countertops, floors and walls of the eight locations noted in the attached diagram. All readings obtained were equivalent to background levels and it was concluded that the area was free of contamination.

INSTRUMENT: Picker 655-186 G-M
CALIBRATED: 4/3/85
CHECK SOURCE READING: 13.5 mR/hr
AREAS SURVEYED: 0.02 mR/hr
WIPE TESTS: 0.02 mR/hr
BACKGROUND: 0.02 mR/hr

All radiation sources have been removed from the area and transferred to Clare Community Hospital, License No. 21-15940-02. All radiation warning signs have been removed and/or obliterated. This area may be considered acceptable from a radiation safety standpoint to be released for use for other purposes as desired.



David Close
Consultant

DC/kl

CONTROL NO. 7 953 0