

MATERIALS LICENSE  
SUPPLEMENTARY SHEET

OFFICIAL RECORD COPY

License Number

06-13504-02

Docket or Reference Number

030-12364

Amendment No. 12

Winsted Memorial Hospital  
115 Spencer Street  
Winsted, Connecticut 06098

In accordance with the letter dated February 17, 1997, License Number 06-13504-02 is hereby terminated.



0/1

Date FEB 22 1997

For the U.S. Nuclear Regulatory Commission  
Original Signed By:  
Michelle Beardsley

By  
Nuclear Materials Safety Branch  
Region I  
King of Prussia, Pennsylvania 19406

9703030373 970222  
PDR ADOCK 03012364  
C PDR



ML 10

FEB 22 1997

E. Cortright Phillips  
Chief Operating Officer  
Winsted Memorial Hospital  
115 Spencer Street  
Winsted, CT 06098

Dear Mr. Phillips:

Please find enclosed Amendment No. 12 terminating License No. 06-13504-02 as requested by your letter dated February 17, 1997.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:  
**Michelle Beardsley**

Michelle R. Beardsley  
Division of Nuclear Materials Safety

License No. 06-13504-02  
Docket No. 030-12364  
Control No. 123884

Enclosure:  
Amendment No. 12

OFFICIAL RECORD COPY

DOCUMENT NAME: R:\WPS\MLTR\L0613504.02

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
NAME	Beardsley	<input checked="" type="checkbox"/> NB					
DATE	02/21/97		02/ /97		02/ /97		02/ /97



... PUTTING OUR HEART IN HEALTH CARE

MS6

J-1

06-13504-02

February 17, 1997

Licensing Assistance Section  
Nuclear Materials Safety Branch  
U.S. Nuclear Regulatory Commission, Region I  
475 Allendale Road  
King of Prussia, PA 19406-1415

In re: Winsted Memorial Hospital, Chapter 7 Case No. 96-23984

Dear Director:

Winsted Memorial Hospital terminated Patient Care on October 25, 1996 and filed for Bankruptcy under Chapter 7 on November 15, 1996. It is our intention to terminate Winsted Memorial Hospital's NRC license.

The close out survey of the Nuclear Medicine Department was done by Yankee Atomic Electric Company and we are enclosing a copy of their letter along with the required information.

If there is any additional information you need please advise us.

Sincerely,

E. Cortright Phillips  
Chief Operating Officer

ECP/gdp  
Enclosure

cc: Barbara L. Hankin, Esquire

OFFICIAL RECORD COPY

ML 10

123884

FEB 20 1997



February 3, 1997  
REG 97-016

Attorney Barbara Hankin  
Bankruptcy Trustee  
c/o Winsted Memorial Hospital  
115 Spencer Street  
Winsted, CT 06098

Dear Atty. Hankin:

The closeout survey of the nuclear medicine department on December 11, 1996 was completed. The following summarizes what was done, and what will be required of you:

1. All the radioactive sources were either transferred to another facility or individual, surveyed to be background, and disposed of with radioactive labeling removed, or returned to their supplier.
2. Waste decaying in storage was surveyed to be background, and released for bio-hazard disposal.
3. The closeout radiation survey of the nuclear medicine lab, camera room, and stress lab indicated that all dose rate readings were background, and no radioactive contamination was present anywhere in either of the above rooms.
4. All radiation warning postings were removed from the nuclear medicine lab, camera room, and stress lab.
5. NRC Form 314 (Certificate of Disposition of Materials) filled out, along with the attached source inventory, leak test results, radioactive source transfer information, and radiation survey sheets.
6. You will be required to sign and date NRC Form 314, and mail it along with the attachments, and a cover letter stating your intent to terminate Winsted Memorial Hospital's NRC license to:


Licensing Assistance Section  
Nuclear Materials Safety Branch  
U.S. Nuclear Regulatory Commission, Region I  
475 Allendale Road  
King of Prussia, PA 19406-1415

7. You will also need to send an additional copy of the information in 6) above along with a cover letter to:

Mr. Frederick P. Scheuritzel, Jr.  
State of Connecticut  
Department of Environmental Protection  
79 Elm Street  
Hartford, CT 06106-5127

If you have any questions regarding the above, please do not hesitate to contact me at 508-568-2746.

Sincerely,

A handwritten signature in cursive script, reading "Stephen Spanos".

Stephen Spanos, Health Physicist  
Radiological Engineering Group  
Environmental Engineering Dept.

/emd

Attachments

Winsted Memorial Hospital  
Source Inventory Transfer List  
NRC License Number 06-13504-02  
Expiration Date April 30, 1998

Source	Type	Serial No.	Activity as of 12/11/96	Transfer/Return to NRC License Number	Comments
Co-57	Vial E	S206056-110	$2.821 \times 10^{-2}$ mCi ( $1.044 \times 10^6$ Bq)	02-28598-01	
Ba-133	Vial E	S358001-17	$1.436 \times 10^2$ $\mu$ Ci ( $5.314 \times 10^6$ Bq)	02-28598-01	
Cs-137	Vial E	S356002-16	$1.756 \times 10^2$ $\mu$ Ci ( $6.498 \times 10^6$ Bq)	06-16624-01	
Co-60	Vial E	319-054-07	$4.085 \times 10^0$ $\mu$ Ci ( $1.511 \times 10^5$ Bq)	02-28598-01	
Co-57	LFOV-Flood	A336	$1.219 \times 10^0$ mCi ( $4.510 \times 10^7$ Bq)	06-16624-01	
Co-57 (A&B)	Spot Source	S289018	$4.068 \times 10^{-2}$ $\mu$ Ci ( $1.505 \times 10^3$ Bq)	N/A	Surveyed to be background, radioactive labeling removed, and disposed of.
Co-57 (Left/Right)	Reference	W03748-3	$2.400 \times 10^1$ $\mu$ Ci ( $8.882 \times 10^5$ Bq)	06-16624-01	
Cs-137	Spot	The Nucleus	1 $\mu$ Ci (Exempt Quantity)	N/A	Stephen Spanos 270 Washington Street, Woburn, MA 01801-2703 Transfer Date: 12/11/96



Winsted Memorial Hospital  
Source Inventory List  
NRC License Number 06-13504-02  
Expiration Date April 30, 1998

Source	Type	Serial No.	Activity/ Calibration Date	Activity as of 12/11/96	Leak Test Results ( $\mu\text{Ci}$ )
Co-57	Vial E	S206056-110	5.5 mCi (4/18/91)	$2.821 \times 10^{-2}$ mCi ( $1.044 \times 10^6$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Ba-133	Vial E	S358001-17	270 $\mu\text{Ci}$ (3/14/87)	$1.436 \times 10^2$ $\mu\text{Ci}$ ( $5.314 \times 10^6$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Cs-137	Vial E	S356002-16	220 $\mu\text{Ci}$ (2/21/87)	$1.756 \times 10^2$ $\mu\text{Ci}$ ( $6.498 \times 10^6$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Co-60	Vial E	319-054-07	53 $\mu\text{Ci}$ (6/15/77)	$4.085 \times 10^0$ $\mu\text{Ci}$ ( $1.511 \times 10^5$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Co-57	LFOV-Flood	A336	5 mCi (6/7/95)	$1.219 \times 10^0$ mCi ( $4.510 \times 10^7$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Co-57 (A&B)	Spot Source	S289018	50 $\mu\text{Ci}$ each (8/88)	$4.068 \times 10^{-2}$ $\mu\text{Ci}$ ( $1.505 \times 10^3$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Co-57 (Left/Right)	Reference	W03748-3	100 $\mu\text{Ci}$ (6/1/95)	$2.400 \times 10^1$ $\mu\text{Ci}$ ( $8.882 \times 10^5$ Bq)	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$
Cs-137	Spot	The Nucleus	1 $\mu\text{Ci}$ (Exempt Quantity)	----	$<1.00 \times 10^{-4}$ $\mu\text{Ci}$

\* Minimum detectable activity =  $1.00 \times 10^{-4}$   $\mu\text{Ci}$  at a 95% confidence level.

Survey By: Steph Spanos

Date: 12/11/96

Counted By: Steph Spanos

Date: 12/11/96



Winsted Memorial Hospital  
Source Inventory Transfer and Receipt List  
NRC License Number 06-13504-02  
Expiration Date April 30, 1998

Sources Transferred to: Johnson Memorial Hospital  
201 Chestnut Hill Road  
Stafford Springs, CT 06076-0860  
NRC License Number 06-16624-01

Source	Type	Serial No.	Activity/ Calibration Date	Activity as of 12/11/96
Cs-137	Vial E	S356002-16	220 $\mu$ Ci (2/21/87)	$1.756 \times 10^2$ $\mu$ Ci ( $6.498 \times 10^6$ Bq)
Co-57	LFOV-Flood	A336	5 mCi (6/7/95)	$1.219 \times 10^0$ mCi ( $4.510 \times 10^7$ Bq)
Co-57 (Left/Right)	Reference	W03748-3	100 $\mu$ Ci (6/1/95)	$2.400 \times 10^1$ $\mu$ Ci ( $8.882 \times 10^5$ Bq)

Receipt Acknowledged By: Michelle A. LaBelle

Title: Nuclear Medicine Technologist; Assistant RSO.

Date: December 11, 1996

Billerica Facility  
331 Treble Cove Road  
North Billerica

1/28/1997

YANKEE ATOMIC ELECTRIC CORP.  
25 RESEARCH DRIVE  
WESTBOROUGH, MA 01581

Dear Customer,

We are in receipt of your returned product, details of which can be found on the attached list. Of the 3 line items returned, 0 have resulted in a credit to your account.

The remaining products did not receive credit for one of the following reasons:

Code	Description
------	-------------

L	NOT A DMPC LOT NUMBER
P	NOT A DMPC PRODUCT CODE

If you have any questions about your return, please contact your Customer Service Representative at (800) 255-1572

Sincerely,  
DMPC Customer Service

1/28/1997

Return Goods Acknowledgement

Page 1

DMPC return # 000903  
Customer # 551840  
3rd Party Ref  
Received on 12/05/1996  
Processed on 12/05/1996

Item Number	Item Description	Lot Number	Exp Date	Qty Received	Qty Authorized	Credit/ Reason
S206	CO-57 DOSE CAL REF SRC - 5MC	S206056110		1.000	1.000	L
S358	BA133 DOSE CAL REF SRC, 250UC	S35800117		1.000	1.000	L
S319		S31905407		1.000		P



>>> End of Acknowledgement <<<

OUTSIDE WALL

PARKING LOT

EXIT

X-ray room  
#2



Survey Meter Calibration Dates:  
Eberline ESP-1 4/25/96  
Ludlum Scaler 1/16/96

PATIENT WAITING  
AREA

X-RAY DEPARTMENT CORRIDOR

Winsted Memorial Hospital  
115 Spencer Street  
Winsted, CT 06098

Survey Meter Used: Eberline ESP-1  
Swipes Counted in: Ludlum Model 2200 Scaler

\* All areas surveyed were background ( $0.02 \frac{mR}{hr}$ )  
All equipment and areas swiped showed  
no contamination.

MAGNETIC  
NORTH

GRID NORTH

1/4" = 1.0'

Stephen Spanos 12/11/96

DRESSING ROOMS


Winsted Memorial Hospital  
115 Spencer Street  
Winsted, CT 06098

Survey Meter Used: Eberline ESP-1

Swipes Counted in: Ludlum Model 2200 Scaler

\* All areas surveyed were background (0.02 mR/hr)  
All equipment and areas swiped showed no contamination.

STRESS LAB (ER)



Stephen Spano  
12/11/96

NORTH



Survey Meter  
Calibration Dates:

Eberline ESP-1 4/25/96  
Ludlum Scaler 1/16/96

(6-95)

10 CFR 30.36(c)(1)(iv)

10 CFR 40.42(c)(1)(iv)

10 CFR 70.38(c)(1)(iv)

## CERTIFICATE OF DISPOSITION OF MATERIALS

INSTRUCTIONS: ALL ITEMS MUST BE COMPLETED -- PRINT OR TYPE  
SEND THE COMPLETED CERTIFICATE TO THE NRC OFFICE SPECIFIED ON THE REVERSE

ESTIMATED BURDEN PER RESPONSE TO COMPLY WITH THIS MANDATORY INFORMATION COLLECTION REQUEST: 30 MINUTES. THIS SUBMITTAL IS USED BY NRC AS PART OF THE BASIS FOR ITS DETERMINATION THAT THE FACILITY HAS BEEN CLEARED OF RADIOACTIVE MATERIAL BEFORE THE FACILITY IS RELEASED FOR UNRESTRICTED USE. FORWARD COMMENTS REGARDING BURDEN ESTIMATE TO THE INFORMATION AND RECORDS MANAGEMENT BRANCH (T-6 P33), U.S. NUCLEAR REGULATORY COMMISSION, WASHINGTON, DC 20555-0001, AND TO THE PAPERWORK REDUCTION PROJECT (3150-0028), OFFICE OF MANAGEMENT AND BUDGET, WASHINGTON, DC 20503. AN AGENCY MAY NOT CONDUCT OR SPONSOR, AND A PERSON IS NOT REQUIRED TO RESPOND TO, A COLLECTION OF INFORMATION UNLESS IT DISPLAYS A CURRENTLY VALID OMB CONTROL NUMBER.

LICENSEE NAME AND ADDRESS

Winsted Memorial Hospital  
115 Spencer Street  
Winsted, CT 06098

LICENSE NUMBER

06-13504-02

LICENSE EXPIRATION DATE

April 30, 1998

## A. MATERIALS DATA (Check one and complete as necessary)

THE LICENSEE OR ANY INDIVIDUAL EXECUTING THIS CERTIFICATE ON BEHALF OF THE LICENSEE CERTIFIES THAT:  
(Check and/or complete the appropriate item(s) below.)

- ☐ 1. NO MATERIALS HAVE EVER BEEN PROCURED OR POSSESSED BY THE LICENSEE UNDER THIS LICENSE.
- OR
- ☒ 2. ALL ACTIVITIES AUTHORIZED BY THE LICENSE HAVE CEASED AND ALL MATERIALS PROCURED AND/OR POSSESSED BY THE LICENSE NUMBER CITED ABOVE HAVE BEEN DISPOSED OF IN THE FOLLOWING MANNER. (If additional space is needed, use the reverse side or provide attachments.)

Describe specific material transfer actions and, if there were radioactive wastes generated in terminating this license, the disposal actions including the disposition of low-level radioactive waste, mixed waste, Greater-than-Class-C waste, and sealed sources, if applicable.

For transfers, specify the date of the transfer, the name of the license recipient, and the recipient's NRC license number or Agreement State name and license number.

See attached sheets

If materials were disposed of directly by the licensee rather than transferred to another licensee, licensed disposal site or waste contractor, describe the specific disposal procedures (e.g., decay in storage)

See attached sheets

## B. OTHER DATA

- ☒ 1. OUR LICENSE HAS NOT YET EXPIRED; PLEASE TERMINATE IT.
2. A RADIATION SURVEY WAS CONDUCTED BY THE LICENSEE TO CONFIRM THE ABSENCE OF LICENSED RADIOACTIVE MATERIALS AND TO DETERMINE WHETHER ANY CONTAMINATION REMAINS ON THE PREMISES COVERED BY THE LICENSE. (Check one)
- ☐ NO (Attach explanation)
- ☒ YES, THE RESULTS (Check one)
- ☒ ARE ATTACHED, or
- ☐ WERE FORWARDED TO NRC ON (Date)

3. THE PERSON TO BE CONTACTED  
REGARDING THE INFORMATION  
PROVIDED ON THIS FORM

NAME

Stephen Spanos

TELEPHONE NUMBER  
(Include Area Code)

508-568-2746

4. MAIL ALL FUTURE CORRESPONDENCE REGARDING THIS LICENSE TO
- Yankee Atomic Electric Company  
580 Main Street  
Bolton, MA 01740-1398

## CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Attorney Barbara Hankin  
Bankruptcy Trustee

SIGNATURE

Barbara L. Hankin  
Bankruptcy Trustee

DATE

2/11/97

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECTS. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTIONS.



FILE CERTIFICATES AS FOLLOWS:

IF YOU ARE A DISTRIBUTOR OF EXEMPT PRODUCTS, SEND TO:

DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY  
OFFICE OF NUCLEAR MATERIAL SAFETY AND SAFEGUARDS  
U.S. NUCLEAR REGULATORY COMMISSION  
WASHINGTON, DC 20555-0001

ALL OTHERS, IF YOU ARE LOCATED IN:

CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE,  
MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW  
JERSEY, NEW YORK, PENNSYLVANIA, RHODE ISLAND, OR  
VERMONT, SEND APPLICATIONS TO:

LICENSING ASSISTANCE SECTION  
NUCLEAR MATERIALS SAFETY BRANCH  
U.S. NUCLEAR REGULATORY COMMISSION, REGION I  
475 ALLENDALE ROAD  
KING OF PRUSSIA, PA 19406-1415

ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI,  
NORTH CAROLINA, PUERTO RICO, SOUTH CAROLINA,  
TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA,  
SEND APPLICATIONS TO:

NUCLEAR MATERIALS SAFETY SECTION  
U.S. NUCLEAR REGULATORY COMMISSION, REGION II  
101 MARIETTA STREET NW, SUITE 2900  
ATLANTA, GA 30323-0189

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI,  
OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

MATERIALS LICENSING SECTION  
U.S. NUCLEAR REGULATORY COMMISSION, REGION III  
801 WARRENVILLE ROAD  
LISLE, IL 60532-4351

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO,  
HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA,  
NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA,  
OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA,  
TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND  
APPLICATIONS TO:

MATERIAL RADIATION PROTECTION SECTION  
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV  
611 RYAN PLAZA DRIVE, SUITE 400  
ARLINGTON, TX 76011-8064

15:11 02 FEB 76

RECEIVED-REGION I



... PUTTING OUR HEART IN HEALTH CARE

06-13504-02  
030-12364

Jl

November 19, 1996

Mr. Thomas T. Martin  
Regional Administrator  
United States Nuclear Regulator Commission  
Region I  
475 Allendale Road  
King of Prussia, PA 19406-1415

Dear Mr. Martin:

This is to officially notify you that effective at 5:00 P.M. on Friday, October 25, 1996 Winsted Memorial Hospital terminated all patient care services.

Subsequent to that date on November 15, 1996, a voluntary petition in bankruptcy was filed with the United States Bankruptcy Court for the district of Connecticut, the Hartford Division.

If you have any questions please give me a call at (860)738-6616.

Sincerely,

E. Cortright Phillips  
Chief Operating Officer  
ECP/gp

U. S. Nuclear Regulatory Commission		Date: 11-18-96
Telephone or Verbal Conversation Record		Time: 10:00 a.m.
<input type="checkbox"/> Incoming Call <input checked="" type="checkbox"/> Outgoing Call <input type="checkbox"/> Visit		
<b>Person Calling:</b> Michelle Beardsley	<b>Office:</b> USNRC Region I	<b>Phone #:</b> (215) 337-6942
<b>Person Called:</b> Stephen Spanos-HP cons.	<b>Office:</b>	<b>Phone #:</b>
<b>Conversation</b>		
<b>Subject:</b> Winstead Memorial Hospital- license termination License No. 06-13504-02; Docket No. 030-12364; Control No. 123884		
<b>Summary:</b> I explained to Mr. Spanos that, in order to terminate the hospital's license, a member of the hospital's management will need to submit the request with the closeout survey and completed form 314.  He stated that the closeout survey is expected to be performed within the next 30 days.		
<b>Referred to:</b>		
<b>Action Requested:</b>  <div style="text-align: center;">Letter</div>		
<b>Action Taken:</b>		

OFFICIAL RECORD COPY

**ML 10**

# YANKEE ATOMIC ELECTRIC COMPANY

Telephone (508) 779-6711  
TWX 710-380-7619  
**030-12364**



580 Main Street, Bolton, Massachusetts 01740-1398

November 7, 1996  
REG 158/96

Dr. Mohamed M. Shanbaky  
Licensing Assistance Team  
Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission, Region I  
475 Allendale Road  
King of Prussia, PA 19406-1415

RE: Winsted Memorial Hospital  
NRC Materials License Termination  
License No. 06-13504-02  
Expiration April 30, 1998

Dear Dr. Shanbaky:

This letter is in regards to Winsted Memorial Hospital, Winsted, CT. The hospital will be filing for bankruptcy and will be terminating their NRC license.

The activities for license termination are listed below:

1. Filing NRC Form 314, Certificate of Disposition of Materials.
2. Leak testing, and transfer of all sealed sources back to their suppliers, or hospital facilities.
3. Final site area radiation and contamination survey results, including survey maps, of the nuclear medicine lab and patient testing areas.
4. Awaiting decay in storage of hospital waste, prior to disposal.
5. Filing appropriate documentation of the above with the NRC and the Connecticut Department of Environmental Protection (DEP).
6. Notification to the NRC under 10CFR30.34(h)(1) and (2) by the licensee.

Dr. Mohamed M. Shanbaky

Page 2

November 7, 1996

7. No nuclear medicine lab equipment, or associated nuclear medicine patient testing equipment, is to be sold or released prior to the close-out survey of the lab and patient testing areas.

Please feel free to contact me at 508-568-2746 with any questions.

Sincerely,



Stephen Spanos  
Health Physicist

SS/111

c: R.A. Marcello  
P.S. Littlefield  
R.D. Cardarelli  
D.B. Everhardt (NRC, Region I)

BETWEEN:

LICENSE FEE MANAGEMENT BRANCH, ARM  
AND  
REGIONAL LICENSING SECTIONS

(FOR LFMS USE)  
INFORMATION FROM LTS

PROGRAM CODE: 02120  
STATUS CODE: 0  
FEE CATEGORY: 7C  
EXP. DATE: 20030430  
FEE COMMENTS:  
DECOM FIN ASSUR REQD: N

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED

APPLICANT/LICENSEE: WINSTED MEMORIAL HOSPITAL  
RECEIVED DATE: 961112  
DOCKET NO: 3012364  
CONTROL NO.: 123884  
LICENSE NO.: 06-13504-02  
ACTION TYPE: TERMINATION

2. FEE ATTACHED

AMOUNT: -----  
CHECK NO.: -----

3. COMMENTS

SIGNED  
DATE

*Rebecca J. Brown*  
11/19/96

B. LICENSE FEE MANAGEMENT BRANCH (CHECK WHEN MILESTONE 03 IS ENTERED 146/96)

1. FEE CATEGORY AND AMOUNT: *7C* *Term*

2. CORRECT FEE PAID. APPLICATION MAY BE PROCESSED FOR:

AMENDMENT -----  
RENEWAL -----  
LICENSE -----

3. OTHER -----

SIGNED  
DATE

-----  
-----

RECEIVED BY LFDOB	
Date	12/9/96
By	Dec 7 I (ST)
By	BB
Date	12/7/96