

AUG 19 1985

LeRoy L. Schroeder, M.D.  
200 W. Pearl Street  
Findlay, OH 45840

Gentlemen:

Enclosed is Amendment No. 02 to your NRC License No. 34-18674-01 in accordance with your request.

Please note that two subitems, barium-133 and cesium-137, have been added to your list of authorized material, along with a leak test condition.

Because you are not authorized for any of Groups I through VI of 10 CFR Part 35.100, you are not subject to the provisions of Section 35.14 of 10 CFR Part 35. This section authorizes group medical licensees to possess sealed sources for calibration of instruments and to perform leak tests on them. Therefore, these license conditions were necessary to give you, a non-group licensee, authorization to possess your barium-133 and cesium-137 sealed sources for calibration of your dose calibrator and to require you to leak test the sources.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.

B509110347 B50816  
REG3 LIC30  
34-18674-01 PDR

AUG 19 1985

6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By  
Cassandra McDonald  
Materials Licensing Section

Enclosure(s): Amendment No. 02

RIII

McDonald/cm,  
08/08/85 8/12/85