

## MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

## Licensee

1. Kalamazoo Orthopaedic Clinic, P.C.

3. License number 21-24536-01

2. 403 Bronson Medical Center, West  
Kalamazoo, MI 49007

4. Expiration date September 30, 1990

5. Docket or  
Reference No. 030-287766. Byproduct, source, and/or  
special nuclear material7. Chemical and/or physical  
form8. Maximum amount that licensee  
may possess at any one time  
under this license

A. Iodine-125

A. Sealed sources  
(AECL Model Nos. C-235  
and C-324 in C-236  
source holder)A. 2 sources not  
to exceed 400  
millicuries each

## 9. Authorized Use

A. One source to be used in a Norland Model N2740 bone densitometer for determination of bone mineral content in humans. One source to be stored in its shipping container for source replacement purposes.

## CONDITIONS

10. Licensed material shall be used only at the licensee's facilities located at 403 Bronson Medical Center, West, Kalamazoo, Michigan.

11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."

12. Licensed material shall be used by William M. Uggen, M.D.

13. A. (1) Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.

(2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.

8509130250 850904  
REG3 LIC30  
21-24536-01

PDR

COPY

2  
ml30

B5 974/85

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number  
21-24536-01

Docket or Reference number  
030-28776

- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.
- D. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. The licensee is authorized to hold radioactive material with a physical half-life of less than 65 days for decay-in-storage before disposal in ordinary trash provided:
- A. Radioactive waste to be disposed of in this manner shall be held for decay a minimum of ten (10) half-lives.
- B. Prior to disposal as normal waste, radioactive waste shall be monitored to determine that its radioactivity cannot be distinguished from background with typical low-level laboratory survey instruments. All radiation labels will be removed or obliterated.
- C. Generator columns shall be segregated so that they may be monitored separately to ensure decay to background levels prior to disposal.
15. Sealed sources containing licensed material shall not be opened or removed from their respective source holders by the licensee.
16. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated July 5, 1985 and ALARA Program received with application dated July 5, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

For the U.S. Nuclear Regulatory Commission

Date September 4, 1985

Original Signed  
By Evelyn R. Matson  
Materials Licensing Section, Region III

COPY