

file

MAY 31 1985

DKN  
03021078

Endress & Hauser, Inc.  
ATTN: Daniel G. Juanillo  
Field Service Manager  
P.O. Box 246-1  
Greenwood, IN 46142

Re: Your Application To Perform Service Operations On Ohmart Corporation Devices

Gentlemen:

As discussed in a telephone conversation between Mr. Daniel G. Juanillo and me on May 30, 1985, I misunderstood the type of license for which you applied in May 1984. Consequently, the information I requested in our February 27, 1985 letter to you is not needed. Instead, we need information on the following items in order to complete our review:

1. It appears that you desire to leak test Ohmart Corporation devices that are in use at various facilities, verify the installation of these devices and initial training of users, and perform tests to assess proper shutter function of these devices. Please confirm if our assumption is correct. If not, please describe each specific service operation you desire to perform and the location(s) where byproduct material will be possessed or used in conjunction with each operation (e.g., customer's facilities).
2. For leak testing, please submit the information requested in Items 5, 7, 8, 9B, 10, and 11 of the enclosed package entitled "Appendix A".
3. For verification of installation or shutter operation, instruction of users, or installation type operations, please submit the information requested in Items 5, 7, 8, 9B, 10, and 11 of the enclosed package entitled "Appendix C".
4. If you will be installing devices that were transferred to individuals under the general license specified in Section 31.5 of 10 CFR Part 31, please describe the steps you will take to ensure the following:
  - a. These licensees have current copies of Section 31.5 of 10 CFR 31, Sections 30.34 and 30.51 through 30.63 of 10 CFR 30, Sections 20.402 and 20.403 of 10 CFR 20, and Appendix D, 10 CFR 20.
  - b. No person could receive more than 0.5 rem in a calendar year under ordinary circumstances of use from any installed generally licensed device.

D. G. Juanillo  
no response  
to any items  
in item 10  
of Appendix A  
i.e. instructions  
on handling  
materials  
Surveyors  
not required  
under item 10

8509130341 850816  
REG3 LIC30  
13-23437-01 PDR

Endress & Hauser, Inc.

2

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 17579.

Sincerely,

Original Signed By  
Bruce S. Mallett  
Materials Licensing Section

Enclosures:

1. Appendix A
2. Appendix C
3. 10 CFR Parts 20, 30 and 31

RIII

*BSM 5/31/85*

Mallett/cm  
05/31/85