

SAFETY INSPECTION

DMB COPY

1. LICENSEE <b>AMOCO OIL COMPANY</b> <b>2815 INDIANAPOLIS BLVD.</b> <b>WHITING, IN 46394</b>		2. REGIONAL OFFICE <b>U.S. NUCLEAR REGULATORY COMMISSION</b> <b>REGION III</b> <b>799 ROOSEVELT ROAD</b> <b>GLEN ELLYN, IL 60137</b>
3. DOCKET NUMBER(S) <b>030-04325</b>	4. LICENSE NUMBER(S) <b>13-00155-10</b>	5. DATE OF INSPECTION <b>8-28-85</b>

Licensee:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission's (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews, with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☐ 1. Within the scope of this inspection, no violations were observed.
- ☒ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- ☒ 3. During this inspection certain of your activities, as checked below, were in violation of NRC requirements.  
**THIS IS A NOTICE OF VIOLATION** which is required to be posted in accordance with 10 CFR 19.11.
- ☐ A. \_\_\_\_\_ was not properly posted to indicate the presence of a \_\_\_\_\_, 10 CFR 20.203(b), (c), (d), (e) or 34.42.
- ☐ B. Containers located in \_\_\_\_\_ were not properly labeled to indicate the presence of radioactive material. 10 CFR 20.203(f)(1), or (f)(2).
- ☐ C. \_\_\_\_\_ of sealed sources were not performed at the proper frequencies. 10 CFR \_\_\_\_\_ License Condition Number \_\_\_\_\_.
- ☐ D. Records of \_\_\_\_\_ were not properly maintained. 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_.
- ☐ E. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
- ☐ F. Reports or notifications of \_\_\_\_\_ were not made in accordance with 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_.
- ☒ H. **FAILURE TO PERFORM INTERNAL INSPECTION OF EACH RADIOGRAPHER & RADIOGRAPHER'S ASSISTANT INVOLVED IN RADIOGRAPHY WORK AT THREE MONTH INTERVALS AS REQUIRED BY 10CFR 34.11(d).**
- ☐ I. \_\_\_\_\_
- ☒ J. **FAILURE TO MAINTAIN DOCUMENTATION ON SHIPPED SEALED SOURCES AS SPECIFIED IN 49 CFR 173.476 AS REQUIRED BY 10CFR 71.5(a).**
- ☐ K. \_\_\_\_\_
- 8509160111 850828  
REG3 LIC30  
13-00155-10 PDR

I hereby state that within 30 days the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

<b>Michael M. Skendrick</b> SIGNATURE - LICENSEE	<b>8-28-85</b> DATE	<b>S. R. Lank</b> SIGNATURE - NRC INSPECTOR	<b>8/28/85</b> DATE
-----------------------------------------------------	------------------------	------------------------------------------------	------------------------

0/1 IE07