

J. POTTER
N51

King's Daughters' Hospital
Staunton, Virginia

Done
8/1/85

Department of Radiology
Paxton P. Powers, M.D.
Kenneth L. Dwyer, M.D.
Dominick M. Conca, M.D.
James B. Huggins, M.D.

July 10, 1985

85 JUL 16 A 9:31

U.S. N.R.C.
LIC. FEE MGMT. BRANCH

85 JUL 22 AM 10:27

RECEIVED

U.S. Nuclear Regulatory Commission
Region II
101 Marietta Street, N.W.
Suite 3100
Atlanta, Georgia 30303

Dear Sir:

Please amend Materials License 45-02367-01 as follows:

This amendment will authorize us to receive the Siemens Gammasonics, Inc., Americium-241 Anatomical Marker, Model SS-10244 and the Americium-241 Dual Isotope Motion Correction Device, Model 035-4230C. Each sealed source will not exceed 14.0 mCi allowing for a total possession limit of 28.0 mCi.

Both of these devices will correct and compensate for the signal output of body movement on a medical Scintillation gamma camera. The Anatomical Marked and Dual Isotope Motion Correction Device have been submitted to NRC and have been accepted for licensing purposes.

Also please delete as authorized users, the following:

Kenneth L. Dwyer, M. D.
Forest G. Harper, M. D.

Enclosed you will find the amendment fee of \$120.00.

Sincerely,

Paul Flanagan
Paul Flanagan
Administrator

James B. Huggins, MD

James B. Huggins, M. D.
Radiologist
Radiation Safety Officer

July - 3 - II

Application	21858	7C	\$120
Check No.			
Amount/ Fee Category			
Type of Fee	Cmd 7/22/85		
Date Check Rec'd	7/22/85		
Receiver	JACQUES		

JBH/cn

Enclosure

8508290320 850805
REG2 LIC30
45-02367-01 PDR

50660
Jed 7/14/85
Official Copy