

NMSS:CLC
Mail Control No. 17574

APR 26 1985

St. Peter's Community Hospital
ATTN: Howard Purcell, Jr.
Executive Director
2475 Broadway
Helena, Montana 59601

Gentlemen:

This refers to your application dated May 22, 1984, to renew your license for use of sealed sources for medical teletherapy. We have completed review of your application and have the following comments and need for additional information:

1. Your application did not list the names of Drs. Boughn, Maxwell, and Fine. We will assume that these physicians, who were previously listed as users under this license, should no longer be listed unless you state otherwise.
2. Confirm that your radiation safety committee will include a member of the institution's management and the Radiation Safety Officer, Dr. Pedersen, as required by 10 CFR Part 35, Section 35.11(b).
3. Confirm that the teletherapy unit has not been relocated since the last survey conducted November 1981. Also identify the room number where the unit is located.
4. Your application requested authorization for a Neutron Products, Inc. Model NPI-20-4500T teletherapy source. We assume that you intended to request Model NPI-20-4500W. According to your most recent survey report, this is the model number of the source last installed in your unit. If our assumption is incorrect, please clarify.
5. Confirm that you have established written operating procedures addressing the topics identified in Appendix G of the attached draft Regulatory Guide, "Guide for the Preparation of Applications for Licenses in Medical Teletherapy Programs" (attached). You need not submit the actual procedures.

Submit a copy of your emergency procedures for review. These should address the topics identified in Item i of Appendix G.

NMSS etc
CLCain;df
4/22/85

C: NMSS
RJEverett
4/25/85

8508300073 850626
REQ4 LIC30
25-12453-03 PDR

6. Identify Health Physics Associates, your identified leak test analyzer, as to address and radioactive materials license number. Also indicate the model number of the leak test kit to be used.

We shall continue review of your application upon receipt of the information and revisions requested above, in duplicate.

Sincerely,

**Original Signed By
C. L. Cain**

C. L. Cain
Nuclear Materials Safety Section

Enclosures: As stated