

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

29-04481-02

Docket or Reference number

030-00354

Amendment No. 12

V. A. Medical Center
Tremont Avenue
East Orange, New Jersey 07019

In accordance with letter dated May 3, 1985, License number 29-04481-02 is amended as follows:

Items 6. 7. 8., and 9. are amended to read:

6. Byproduct, source, and/or special nuclear material

A. Cobalt 60

7. Chemical and/or physical form

A. Teletherapy sealed sources (Picker Model P3801A, Advanced Medical Systems, Inc. Model AMS-3801, or Neutron Products, Inc., Model NPI-15-6000W)

8. Maximum amount that licensee may possess at any one time under this license

A. 15,200 curies (2 sources of not more than 7,600 curies each)

9. Authorized use

A. One source to be used in a Picker Model 6296 (C9M/80) teletherapy unit for the treatment of humans and irradiation of mammals. One source in its shipping container to be in possession of the licensee as necessary to the replacement of the source in the teletherapy unit only.

8508300373 850722
REG1 LIC30
29-04481-02 PDR

For the U.S. Nuclear Regulatory Commission

Original Signed By:

By

Jerome M. Johansen

Nuclear Materials Safety and
Safeguards Branch, Region I
King of Prussia, Pennsylvania 19406

Date JUL 22 1985

"OFFICIAL RECORD COPY"

ML10