

MATERIALS LICENSE  
SUPPLEMENTARY SHEET

License number

34-15317-01

Docket or Reference number

030-08906

Amendment No. 07

Shelby Memorial Hospital  
Morris Road  
Shelby, OH 44875

In accordance with letter dated July 31, 1985, License Number 34-15317-01 is amended as follows:

Items 6., 7., 8., and 9. are amended to add:

6. Byproduct, source,  
and/or special nuclear  
material

7. Chemical and/or  
physical form

8. Maximum amount that  
licensee may possess  
at any one time  
under this license

D. Xenon-133

D. Gas or gas in solution  
that is the subject  
of an active (i.e., not  
withdrawn or terminated)  
"New Drug Application"  
(NDA) approved by FDA or  
an active (i.e., not  
withdrawn, terminated  
or on "clinical hold")  
"Notice of Claimed  
Investigational Exemption  
for a New Drug" (IND)  
that has been accepted  
by FDA

D. 100 millicuries

9. Authorized Use

D. Blood flow studies. Pulmonary function studies.

Condition 19. is amended to read:

19. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated November 15, 1982; letters dated October 25, 1982 and July 31, 1985; and ALARA Program dated October 24, 1980. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

8509120016 850829  
REG3 LIC30  
34-15317-01 PDR

COPY

2  
meze

BS 8/30/85

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number

34-15317-01

Docket or Reference number

030-08906

Amendment No. 07

Condition 20. is added:

20. The licensee shall measure the airflow rates initially and semi-annually thereafter to assure proper ventilation system performance. Records of the results of ventilation measurements, instrument used to perform measurements, instrument calibration data and person performing the airflow rate tests, shall be maintained.



For the U.S. Nuclear Regulatory Commission

Original Signed

By Patricia J. Whiston

Materials Licensing Section, Region III

Date August 29, 1985

COPY