

DS09

1/0

Received Rules & Procedures
Branch - 1/17/86

The University of Texas Medical Branch at Galveston



77550

Medical School
Graduate School of Biomedical Sciences
School of Allied Health Sciences
School of Nursing

Marine Biomedical Institute
Institute for the Medical Humanities
UTMB Hospitals at Galveston

Published

12/30/85

50FR53226

NUCLEAR MEDICINE DIVISION
RADIOLOGY DEPARTMENT
(409) 761-2921

ANTHONY R. BENEDETTO, PH.D.
ASSOCIATE PROFESSOR

January 10, 1986

86-2

SUBJECT: Task FC 414-4

Pat Vacca
Rules and Procedures Branch
DRR, ADM
US Nuclear Regulatory Commission
Washington, DC 20555

Dear Pat:

I have reviewed the December 1985 draft of the subject document and have only two comments.

1. On pages 4 and 5, the current Agreement States could be indicated by bold face type or underlining. The footnote could then state that an applicant should check with NRC if his state was not indicated as being an Agreement State.
2. On page 24 you might wish to consider a discussion of acceptable instrumentation and calibration procedures for the beam intensity measuring instruments. They are such an important part of the program that it seems strange that they are not specifically addressed.

Sincerely,

Anthony R. Benedetto, Ph.D.

ARB/skb


BK6 - 1/21/86

8601240291 860110
PDR REGGD
10. XXX C PDR

NUCLEAR MEDICINE DIVISION
THE UNIVERSITY OF TEXAS
MEDICAL BRANCH
GALVESTON, TEXAS 77550

STATE OF TEXAS
OFFICIAL BUSINESS
PENALTY FOR
PRIVATE USE

1111
1111
1111
1111



Pat Vacca
Rules and Procedures Branch
DRR, ADM
US Nuclear Regulatory Commission
Washington, DC 20555

