

File

NOV 12 1985

Procter & Gamble Manufacturing Company
ATTN: Michael Kenron
1232 West North Avenue
Chicago, IL 60622

Enclosed is Amendment Number 12 to your NRC License 12-12842-01 which authorizes the changes you requested in your letter dated October 15, 1985.

Please note that the Cadmium-109 sealed source model no. NER-465 has not been evaluated or approved by the NRC for a leak test frequency in excess of 6 months. In order to be evaluated for an extended leak test frequency you or the manufacturer must request that the sealed source be reevaluated for the new frequency. For a description of the information that the NRC needs refer to the enclosed guidelines.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

Sincerely,

Original Signed By
Evelyn R. Matson
Materials Licensing Section

8512030370 851112
REG3 LIC30
12-12842-01 PDR

RILL
EM
Matson/cm
10/30/85