

FEB 12 1985

Bemidji State University
ATTN: Dr. Gary W. Evans
Associate Professor of Chemistry
Division of Science and Math
Bemidji, Minnesota 56601

Gentlemen:

Enclosed is Check 404284 which accompanied your January 16, 1985 application for renewal of Materials License 22-07944-01.

Section 170.11(a)(9) of Part 170, copy enclosed, provides that no fees will be required for "A license for possession and use of byproduct material, source material, or special nuclear material applied for by, or issued to an agency of a State or any political subdivision thereof." The Bemidji State University is, therefore, exempt from payment of license fees.

Your application has been sent to the Licensing staff for processing.

Sincerely,

Original Signed by
A. S. Cabell

for

Glenda Jackson
License Fee Management Branch
Office of Administration

Enclosures:

1. Check 404284 (\$150)
2. 10 CFR 170

DISTRIBUTION:

Pending Fee File
Weekly Reading File
Materials Reading File

8512030541 851025
REG3 LIC30
22-07944-01 PDR

OFFICE	LFMB:ADM	LFMB:ADM					
SURNAME	CPhillips:rej	GJackson					
DATE	2/11/85	2/11/85					

CONVERSATION RECORD

TIME

1100 AM

DATE

10/4/85

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Dr. Lindgren

ORGANIZATION (Office, dept., bureau, etc.)

Bemidji State U

TELEPHONE NO.

218-755-

2920

2798

SUBJECT

LIC NO 22-07944-01 Renewal

Applic dated 1/16/85 & letter dated 9/27/85

SUMMARY

1. Leak test procedures for PuBe neutron source as described in letter dated 9/27/85 does not appear to be an adequate evaluation. The Calibration source of Pb-210 has undergone 34 half-lives since it was documented as 0.1 μ Ci in 1972. The current activity of 3×10^{-12} μ Ci is unmeasurably small.
2. Describe the window on the electroscope.
3. Describe the geometry for measuring standard & wipe.
4. Morine must obtain or document training and experience with radionuclides that he wishes to use. He currently shows no training in basic radioisotope handling techniques and has experience with Co-60 sealed source and a chemical process involving ^{radio-}Iodine

ACTION REQUIRED

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

ACTION TAKEN

SIGNATURE

TITLE

DATE