

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number 21-02335-08

Docket or Reference number 030-08659

Amendment No. 06

The Detroit Edison Company
Generation Engineering
2000 Second Avenue
Detroit, MI 48226

In accordance with letter dated October 17, 1985, License Number 21-02335-08 is amended as follows:

Items 6., 7., 8., and 9. are amended to add:

6. Byproduct, source, and/or special nuclear material

7. Chemical and/or physical form

8. Maximum amount that licensee may possess at any one time under this license

G. Americium-241

G. Sealed source (Troxler Dwg. No. A-102451)

G. No single source to exceed 50 millicuries

9. Authorized Use

G. To be used in Troxler Model No. 3216 moisture gauge to measure properties of construction materials.

Condition 12. is amended to read:

12. A. Licensed material shall be used by, or under the supervision and in the physical presence of, individuals who have satisfactorily completed the device manufacturer's training program for gauge users and have been designated by the licensee's Radiation Protection Officer. The licensee shall maintain records of the individuals who have been designated as authorized users.

B. The Radiation Protection Officer for the activities authorized by this license is Robert V. Collins.

8512040046 851107
REQ3 LIC30
21-02335-08 PDR

For the U.S. Nuclear Regulatory Commission

NOV 7 - 1985

Date _____

Original Signed
By B. J. Holt
Materials Licensing Section, Region III

ML30 Bx NOV 7 - 1985 COPY 2