

MATERIALS LICENSE  
SUPPLEMENTARY SHEET

License number

37-21421-01

Docket or Reference number

030-20636

Amendment No. 01

"OFFICIAL RECORD COPY"

City of Bethlehem  
Water Laboratory  
10 E. Church Street  
Bethlehem, Pennsylvania 18018

In accordance with letter dated August 14, 1985, License Number 37-21421-01 is amended as follows:

Condition 14.A. is amended to read:

14. A. Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed three years. In the absence of a certificate from a transferor indicating that a test has been made within three years prior to the transfer, a detector received from another person shall not be put into use until tested.



For the U.S. Nuclear Regulatory Commission

Original Signed By:

By John D. Kinneman

Date SEP 10 1985

Nuclear Materials Safety and  
Safeguards Branch, Region I  
King of Prussia, Pennsylvania 19406

ML10

8509190043 850910  
REG1 LIC30  
37-21421-01 PDR