

# City of Bethlehem

INDUSTRY • EDUCATION • RELIGION • MUSIC • RECREATION  
AN EQUAL OPPORTUNITY EMPLOYER

DEPARTMENT OF PUBLIC WORKS

August 14, 1985

U.S. Nuclear Regulatory Commission  
Region I  
Office of Inspection and Enforcement  
631 Park Avenue  
King of Prussia, PA 19406-1498

Re: License number 37-21421-01  
Docket or Reference number 030-20636  
City of Bethlehem Water Laboratory  
10 East Church Street  
Bethlehem, Pennsylvania 18018

Dear Sir:

I am requesting an amendment to the above referenced license with regards to item 14.A. such that item 14.A. be changed to:

14.A. Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed 36 months. In the absence of a certificate from a transferor indicating that a test has been made within 36 months prior to the transfer, a detector received from another person shall not be put into use until tested.

Enclosed is a notice from Tracor Instruments stating that their general and specific licenses have been amended to show testing every 36 months for their Nickel 63 electron capture detectors.

If you have any questions, please call me at 215-865-7144.

Sincerely,

*Jeffrey A. Andrews*  
Jeffrey A. Andrews  
Chemist

*Aug - 22 - I*

Applicant	.....
Check No.	.....
Amount	.....
Type of Fee	.....
Date Check Rec'd	.....
Received By	<i>J. C. Anspach</i>

Enclosure  
JAA:cm  
cc: J.C. Anspach  
Lab  
File

8509190060 850910  
REG1 LIC30  
37-21421-01 PDR



"OFFICIAL RECORD COPY"

ML10

04263

FEE EXEMPT 170.11(a)(19)

10 EAST CHURCH STREET, BETHLEHEM, PENNSYLVANIA 18018

AUG 15 1985

\*\*\*\*\*  
N O T I C E  
\*\*\*\*\*

Effective November 7, 1983, all Tracor Ni63 electron capture detectors will require wipe testing only once every three years (36 months). In order to take advantage of this new interval, your specific license must be amended to show testing every 36 months. You may find your state or jurisdiction will require more frequent testing.

This extended interval is covered under Amendment 23 of our Specific License 6-1186 and Amendment 1 of our General License 6-3389GL.

Should you have any additional questions or require assistance with this new benefit, please contact our Radiation Officer, Mr. Carlton Lorfing, at (512) 926-2800, Ext. 5013.

/ve

Amended on April 24, 1984, per The State of Texas.

BETWEEN: William O. Miller, Chief  
License Fee Management Branch  
Office of Administration

John E. Glenn, Chief  
Nuclear Materials Section B  
Division of Engineering and  
Technical Programs

LICENSE FEE TRANSMITTAL

A. REGION I Fee Exempt

1. APPLICATION ATTACHED

Applicant/Licensee: City of Bethlehem

Application Dated: 8/14/85

Control No.: 04263

License No.: 37-21421-01

2. FEE ATTACHED

Amount: 0

Check No.: 0

3. COMMENTS

Signed Brenda P. Latchek

Date 8/20/85

03123  
~~03320~~

B. LICENSE FEE MANAGEMENT BRANCH

1. Fee Category and Amount: ~~30~~ EX 3P 8/ps

2. Correct Fee Paid. Application may be processed for:

Amendment ✓

Renewal \_\_\_\_\_

License \_\_\_\_\_

Signed Rita Jacques / ps

Date 9/3/85