

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

06-15298-02

Docket or Reference number

030-21256

Amendment No. 01

"OFFICIAL RECORD COPY"

Stamford Health Department Laboratory
229 North Street
Stamford, Connecticut 06902

In accordance with letter dated July 3, 1985, License Number 06-15298-02 is amended as follows:

Conditions 14. and 17. are amended to read:

14. A. (1) Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a detector received from another person shall not be put into use until tested.
- (2) The periodic leak test required by this condition does not apply to detector that are stored and not being used. The detectors excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within six months prior to the date of use or transfer.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the surfaces of the device in the which the foil is mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the foil from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with U. S. Nuclear Regulatory Commission, Region I, 631 Park Avenue, King of Prussia, Pennsylvania 19406, describing the equipment involved, the test results, and the corrective action taken.
- D. Tests for leakage and/or contamination shall be performed by the manufacturer or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated May 6, 1985, and letters dated June 3, 1985, and July 3, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

8509190137 850815
REG1 LIC30
06-15298-02 PDR

For the U.S. Nuclear Regulatory Commission

Original Signed By:

John E. Glenn

By

Nuclear Materials Safety and
Safeguards Branch, Region I
King of Prussia, Pennsylvania 19406

Date

AUG 15 1985

ML10