



Newton-Wellesley
Hospital

An Affiliate of NeWell Health Corporation
2014 Washington Street
Newton, MA 02162
(617) 964-2800

April 23, 1985

MS 16
P3

John J. Miller
United States Nuclear Regulatory Commission
631 Park Avenue
King of Prussia, Pennsylvania 19406

License No. 20-02615-01
Docket No. 030-01848
Control No. 18723

Dear Mr. Miller,

As requested in Dr. Glenn's letter of April 5, 1985 the Newton-Wellesley Hospital is submitting in duplicate a comprehensive application for renewal of the NRC License. Preceptor statements for individual users, and qualifications for radiation safety officer are on file with NRC.

Sincerely,

Jeffrey Kelly
Senior Vice President and
Chief Operating Officer

JK/kdb

8509190218 850906
REG1 LIC30
20-02615-01 PDR

18723

"OFFICIAL RECORD COPY"

ML10

APR 26 1985



DOSE CALIBRATOR

Our dose calibrator is checked annually by Mr. Neil Gaeta, 35 Grove Street, Medford, MA 02155 (NRC License 20-20743-01), and our survey instruments are calibrated annually by Mr. Gaeta. Survey instruments are sent to Nuclear Instruments Co., Rockland, MA (License 20-10972-01) for repair and recalibration.

Our group carried out dose calibrator measurements as indicated in Appendix D. The following sources are used:

57 Co (NEN) NES-206	5.1 mCi 2/23/85 $\pm 4\%$
137 Cs (NEN) NES-356	202 mCi 11/2/79 $\pm 4.3\%$
133 Ba (NEN) NES-358	272 mCi 1/18/79 $\pm 4\%$
60 Co (NEN) NES-354	50 mCi 10/25/79 $\pm 4\%$

These sources are all trackable to National Bureau of Standards.