

NOV 28 1983

License No. 07-19576-01
Docket No. 030-17902
Control No. 01827

Delaware Department of Agriculture
Regulatory Chemical Laboratory
ATTN: W. H. Garner
Chemist In Charge
Drawer D
Dover, Delaware 19903

Gentlemen:

Enclosed is Amendment No. 01 to License No. 07-19576-01. You should review this amendment for correctness and to assure that any changes in procedures required by the conditions are implemented.

Please note that the six month wipe test need not be performed on the detector cells while in storage. However, you will need to leak test (prior to use) each detector cell you take out of storage to use if the due date on the detector cell for leak testing has passed.

We wish you continued success with your licensed program.

Sincerely,

Original Signed By
John E. Glenn, Ph.D.

John E. Glenn, Ph.D., Chief
Nuclear Materials Section B
Division of Engineering and
Technical Programs

Enclosures:

1. Amendment No. 01
2. Form NRC-313(I)
3. Form NRC-3

8509190556 830828
REG1 LIC30 PDR
07-19576-01

RI:DETP
Varela/lp
11/21/83

RI:DETP
Glenn
11/18/83

OFFICIAL RECORD COPY

200VARELA11/20/83 - 0001.0.0
11/20/83

ML10