



3034263

Robert J. Rodgers, D.O. • James D. Goettsch, D.O.  
650 Youngstown Warren Rd. • Suite C • P.O. Box 28 • Niles, Ohio 44446  
(330) 544-3636 Fax (330) 544-9449 800-246-2976

April 4, 1997

U. S. Nuclear Regulatory Commission  
ATTN: DOCUMENT CONTROL DESK  
Washington, D. C. 20555

Re: Reply to a Notice of Violation  
License No. 34-26762-01

Dear Sir:

This is in reply to your letter of March 20, 1997 in which the results of the safety inspection conducted on February 27, 1997 were stated.

Radiation surveys were not conducted at the end of each day of use in areas where radiopharmaceuticals are routinely prepared for use or administered. This was an oversight in the start-up radiation program for a new licensee. The nuclear medicine technologist initiated daily surveys when informed of the requirement by the inspector. The results of daily surveys will be reviewed by the radiation safety officer or consulting physicist. Full compliance was achieved on February 27, 1997.

Sincerely,

*James D. Goettsch, D.O.*  
James D. Goettsch, D.O.

cc: Regional Administrator  
U. S. Nuclear Regulatory Commission  
Region III  
801 Warrenville Rd.  
Lisle, Illinois 60532 4351

9704090196 970404  
PDR ADOCK 03034263  
C PDR

090041

