

MS-16
K9

HAZLETON RADIOLOGY ASSOCIATES

EDGAR L. DESSEN, M.D.
ARTHUR S. HABER, M.D.
JOSEPH D. KALOWSKY, D.O.
ROMEO C. OUANO, M.D.
HERMENEGILDO H. ALMARIA, JR., M.D.
STANLEY P. LAUCKS, JR., M.D.
RAULIE RODRIGO, M.D.

A PROFESSIONAL CORPORATION
NORTHEASTERN BUILDING
HAZLETON, PENNSYLVANIA 18201
717 - 455-3908
RADIOLOGY

July 18, 1985

John D. Kinneman, Chief
Nuclear Materials Safety Section A
Division of Radiation Safety and Safeguards
U.S. Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, PA 19406


Dear Mr. Kinneman:

This is in response to your request for additional information as you review our application for a byproduct material license. We make reference to your mail Control Number 03945.

In reference to Item 10, please see our emergency procedures, (attached), which will be posted with our DP3 Spine/Femur Scanner.

Thank you for your time and consideration and we will await your response.

Sincerely yours,


Romeo C. Ouano, M.D.

RCO/ja

8509190634 850823
REG1 LIC30
37-20861-01 PDR

12 2 11 02 100 507
HAZLETON - PENNSYLVANIA

"OFFICIAL RECORD COPY"

ML10

103945
JUL 23 1985

HAZLETON RADIOLOGY ASSOCIATES

EDGAR L. DESSEN, M.D.
ARTHUR S. HABER, M.D.
JOSEPH D. KALOWSKY, D.O.
ROMEO C. OUANO, M.D.
HERMENEGILDO H. ALMARIA, JR., M.D.
STANLEY P. LAUCKS, JR., M.D.
RAULIE RODRIGO, M.D.

A PROFESSIONAL CORPORATION
NORTHEASTERN BUILDING
HAZLETON, PENNSYLVANIA 18201
717 - 455-3600
RADIOLOGY

EMERGENCY PROCEDURE

- I. If there is any reason to suspect that the source shutter is stuck in the open position:
 1. All patients and office personnel are to leave the room immediately and a 0.5 mm lead radiology apron placed over the source. Lock the door.
 2. Call the Radiation Safety Office and advise him of the situation.
 3. Contact the Lunar Radiation Corporation and advise them of the situation.
 4. Do not allow access to the room until approval is given by the Radiation Safety Officer.
- II. If there is reason to suspect that the integrity of the sealed source is in question or during an accident in which a disruption of the source may have occurred.
 1. All patient and office personnel are to leave the room immediately.
 2. Call the Radiation Safety Officer and advise him of the situation. Lunar Radiation Corporation will be called immediately.
 3. All windows shall be closed, fans and air conditioners shall be shut off to prevent airborne spread of contamination.
 4. The doors will be closed and locked.
 5. No immediate attempt will be made to clear up the spill. Clean-up will be performed only under the direction of the RSO.
 6. Persons suspected of being contaminated shall be tested for radioactivity and immediate steps taken to remove contamination under the RSO's supervision.
 7. Entrance to the area will not be permitted until authorized by the RSO.

HAZLETON RADIOLOGY ASSOCIATES

EDGAR L. DESSEN, M.D.
ARTHUR S. HABER, M.D.
JOSEPH D. KALOWSKY, D.O.
ROMEO C. OUANO, M.D.
HERMENEGILDO H. ALMARIA, JR., M.D.
STANLEY P. LAUCKS, JR., M.D.
RAULIE RODRIGO, M.D.

A PROFESSIONAL CORPORATION
NORTHEASTERN BUILDING
HAZLETON, PENNSYLVANIA 18201
717 - 455-3608
RADIOLOGY

III. The Radiation Safety Officer (RSO) is:

Romeo C. Ouano, M.D.

Office Phone: (717)-455-3608

Home Phone: (717)-455-4602

Lunar Radiation Corporation: (608)-258-8545