

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License number

12-24537-01

Docket or Reference number

030-28777

Amendment No. 01

Vollmer Associates  
2 Eastport Plaza Drive  
Collinsville, IL 62234

In accordance with letter dated October 15, 1985, License Number 12-24537-01 is amended as follows:

Items 6., 7., 8. and 9. are amended to add:

6. Byproduct, source,  
and/or special nuclear  
material

7. Chemical and/or  
physical form

8. Maximum amount that  
licensee may possess  
at any one time  
under this license

C. Cesium-137

C. Sealed source  
(Campbell Pacific Nuclear  
Model No. 131)

C. No single source  
to exceed 10  
millicuries

D. Americium-241

D. Sealed source  
(Campbell Pacific Nuclear  
Model No. 131)

D. No single source  
to exceed 50  
millicuries

9. Authorized Use

C. and D. To be used in Campbell Pacific Nuclear MC Series moisture/density gauge to measure properties of construction material.

Condition 12. is amended to read:

12. A. Licensed material shall be used by, or under the supervision and in the physical presence of, individuals who have satisfactorily completed the device manufacturer's training program for gauge users and have been designated by the licensee's Radiation Protection Officer. The licensee shall maintain records of the individuals who have been designated as authorized users.

B. The Radiation Protection Officer for the activities authorized by this license is Douglas W. Keene.

8512050137 851107  
REG2 LIC30  
12-24537-01 PDR

For the U.S. Nuclear Regulatory Commission

Date NOV 7 1985

Original Signed  
By B. J. Holt  
Materials Licensing Section, Region III

COPY