

FEB 02 1986

Docket No. 30-02940

License No. 37-00147-02

Paul S. Friedman, M.D.
1422 Chestnut Street
Philadelphia, Pa 19102

Dear Dr. Friedman:

Subject: Inspection No. 85-01

This refers to your letter dated December 9, 1985, in response to our letter dated November 20, 1985.

Thank you for informing us of the corrective and preventive actions documented in your letter. Regarding your request to use other NRC approved calibration services to calibrate your survey meter, you are advised that you must apply for an amendment to your license to reflect this change. This request should be sent, with the appropriate amendment fee, to the USNRC, Nuclear Materials Safety and Safeguards Branch, 631 Park Avenue, King of Prussia, PA 19406. These actions will be examined during a future inspection of your licensed program.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:
John D. Kinneman

John D. Kinneman, Chief
Nuclear Materials Safety Section A,
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
Commonwealth of Pennsylvania

bcc:
Region I Docket Room (w/concurrences)
T. Darden, DRSS

RI:DRSS
Darden/rbk
1/30/86
J.N.D.

RI:DRSS
Kinneman
1/30/86

OFFICIAL RECORD COPY

RL P S FRIEDMAN - 0001.0.0
01/23/86

8602070157 860202
REG1 LIC30
37-00147-02 PDR

RETURN ORIGINAL TO
REGION I

IE:07 1/1