



John R. Inglish

cc: G.J. Trebilcox
E.A. Sonnet

PHYSIDAL INVENTORY
FOILS & SEALED SOURCES

H.J. HEINZ COMPANY
1062 PROGRESS STREET
PITTSBURGH, PA 15212

<u>MATERIAL</u>	<u>HOW USED</u>	<u>LOCATION</u>	<u>CHECKED</u>
Nickel 63	Detector Cell Perkin-Elmer Chromatograph	Research Building #220 Third Floor	<u>JRH</u> 10/10/85 3/10/86 10/10/86
Cesium 137	Texas Nuclear Scale - Coal	Power Building #420 First Floor	<u>JRH</u> 10/10/85 3/10/86 10/10/86
Americium 241	PECO Control Fill Height	Power Building #420 Sixth Floor	<u>JRH</u> 10/10/85 3/10/86 10/10/86
Cesium 137	Kay-Ray Level Control	Mechanical Building #110 Second Floor	<u>JRH</u> 10/10/86 3/10/86 10/10/86

JRI/smw
10/10/85



516 West Campus Drive Arlington Heights, Illinois 60004
Phone (312) 259-5600 Cable Address KAYRAY Telex 78 2536

LEAK TEST CERTIFICATE

To: H.J. Heinz Company
P.O. Box 57
Pittsburgh, PA 15230

Attn: Mr. Larry Glienn

Date:

Ref: BFS-685, Kay-Ray P.O.# 24719

KR Job No: 4744-022

This certifies that the source(s) listed below have been leak tested according to prevailing NRC standards, and radioactive contamination found to be less than .005 μ Ci AmBe 241.

Please retain this certificate for your files.

CERTIFICATION:

By: *William B. Runkin* 125
Title: *Health Physicist*
Date: *11/27/85*

.....

<u>Leak Test Serial No.</u>	<u>Source Holder Manufacturer</u>	<u>Source Holder Model No.</u>	<u>Source Holder Serial No.</u>	<u>Activity (μCi)</u>	<u>Date</u>	<u>By</u>
30822	Peco Controls	GAM-101	5478	100	11/85	F.Cappelli

ARCADIA CONTROLS, INC.

9800 McKnight Road
Pittsburgh, Pennsylvania 15237
Phone: 931-0111 366-4111

Service Number 111485F

Original Job Number

INVOICED TO

Contact
Phone
Customer P.O.

EQUIPMENT SERVICED

PECO CONTROLS

GAMMA DEVICE

REASON FOR SERVICE

Installation Repair
Calibration ☒ Other
LEAK TEST

Service Man F. CAPPELLI

JOB LOCATION

H.J. HEINZ

1062 Progress St.

PITTSBURGH, PA

Contact LARRY GLENN
Phone 237-5894

DATE OF SERVICE

(including travel)

11-14-85 To 11-14-85

To

Comments: PERFORMED LEAK TEST ON LINE 1 FILL
LEVEL DEVICE. PECO MODEL GAM-101, 6TH FLOOR
POWER BLDG.

NOTE: Will notify JOHN INGRAM of the time
requirements for leak tests.

Dennis Volney

Customer Signature/Title

11-14-85

Date

F. Capelli

Service Man



KAY-RAY INC.

INDUSTRIAL PROCESS CONTROL EQUIPMENT

516 West Campus Drive Arlington Heights Illinois 60004
Phone (312)259 5600 Cable Address KAYRAY Telex 28 2536

LEAK TEST CERTIFICATE

To: H.J. Heinz Company
1062 Progress Street
Pittsburgh, PA 15212

Attn: Larry Gleinn

Date: October 30, 1985

Ref: 423712

KR Job No: 4744-020

This certifies that the source(s) listed below have been leak tested according to prevailing NRC standards, and radioactive contamination found to be less than .005 μ Ci CS

Please retain this certificate for your files.

CERTIFICATION:

By: *T. Baker*

Title: Field Engineering Services

Date: October 30, 1985

<u>Leak Test Serial No.</u>	<u>Source Holder Manufacturer</u>	<u>Source Holder Model No.</u>	<u>Source Holder Serial No.</u>	<u>Activity (mCi)</u>	<u>Date</u>	<u>By</u>
28363	T.N.	5094	B204	200	10/85	A.C.Amsler

ARCADIA CONTROLS, INC.

9800 McKnight Road
Pittsburgh, Pennsylvania 15237
Phone: 931-0111 366-4111

Service Number 102585 Act

Original Job Number 13FS-683

INVOICED TO

Contact _____
Phone _____
Customer P.O. 423712

JOB LOCATION

H. J. Heinz Co.

Contact Larry Gleim
Phone 237-5894

EQUIPMENT SERVICED

REASON FOR SERVICE

____ Installation _____ Repair
____ Calibration _____ Other
Wipe / Leak Test

Service Man A. C. Ansh

DATE OF SERVICE

(including travel)

10/25/85 To 10/25/85
____ To _____

Comments: (1) Checked TW Weigh Scale visually
(2) Surveyed scale - OK
(3) Performed wipe test - OK less than .003

L. Gleim 10/25/85
Customer Signature/Title

10/25/85
Date

A. C. Ansh
Service Man