



Hospital of
Saint Raphael
New Haven, Connecticut 06511

January 7, 1986

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards
United States Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, PA 19406

RE: NOTICE OF VIOLATION
Docket No. 030-01238
License No. 06-00200-03

Dear Dr. Glenn:

In response to your letter of December 31, 1985, the Department of Quality Assurance and Utilization Review at the Hospital of St. Raphael will, on a quarterly basis, determine that the Radiation Safety Committee has assured compliance with NRC rules and regulations. This additional management control has been implemented as of December 6, 1985.

In response to the specific violations as noted in Appendix A, the following steps have been taken:


- A. Potentially reportable events have been added to the agenda of Radiation Safety meetings. We have been in compliance since December 6, 1985.
- B2. Dose calibrator accuracy will be tested annually. We have been in compliance since November 14, 1985.
- B3. Survey instruments have been spot-checked and will have full calibration before the due date. We expect to be in full compliance by May 1, 1986.
- B4. Wipe tests are measured in a calibrated Sodium Iodide Well Counter. We have been in compliance since November 7, 1985.
- B5. The Hospital of St. Raphael requires written requests for the therapeutic administration of radiopharmaceuticals. We have been in compliance since May 30, 1985.

8602070273 860202
REG1 LIC30
06-00200-03 PDR

JAN 13 1986

D. Wipe tests have been performed, and the results entered into a log book. We have been in compliance since November 8, 1985.

Sincerely,

A handwritten signature in cursive script, reading "Edward K. Prokop".

Edward K. Prokop, M. D.
Director
Radiation Safety Committee

EKP/ap

cc Ms. Diana Ballard
Mr. Steve Kimball