



UNITED STATES
NUCLEAR REGULATORY COMMISSION
WASHINGTON, D. C. 20555

September 9, 1983

Walker

MEMORANDUM FOR: Richard E. Cunningham, Director
Division of Fuel Cycle and Material Safety
Office of Nuclear Material Safety and Safeguards

FROM: Richard C. DeYoung, Director
Office of Inspection and Enforcement

SUBJECT: REVISION OF 10 CFR PART 35

We have reviewed the subject revision transmitted by your memo dated August 26, 1983 and we have no comments. Harriet Karagiannis, phone number 492-9655, will be the IE contact with the task force members revising 10 CFR Part 35.

A handwritten signature in cursive script, reading "R C DeYoung", is positioned above the typed name and title.

Richard C. DeYoung, Director
Office of Inspection and Enforcement

CONTACT: Harriet Karagiannis, IE
492-9655

8509230633 850906
PDR PR
35 50FR30616 PDR

	<u>Topic</u>	<u>Current</u>	<u>Proposed</u>	<u>Why the Change was Made</u>
A	teletherapy AU T&E	draft tele- therapy RG App B	35.960	require "adequate"
B	teletherapy safety procedures	draft tele- therapy RG App G	35.610, .633	provides workers with a uniform course of action
A	supervision of physicians	RG 10.8 §4.4*	35.38	allows training of physicians and technicians
A	licensure of unnamed users	policy	35.2	legalizes receipt, use, and disposal by technicians
A	RSO T&E	RG 10.8 §4.8	35.900	identifies RSO T&E
A	AU T&E	RG 10.8 App A	35.910, .920 .930, .940	AU needs to know when to prescribe, how to handle safely
A	Survey meter calibration	RG 10.8 App D 1	35.951	survey instruments are needed to measure radiation levels
A	dose calibrator cali- bration	RG 10.8 App D 2	35.950	needed to measure dosages, avoid misadministration
A	area survey procedures	RG 10.8 App I	35.70	assures control, notice of unnecessarily high levels
A	waste disposal	RG 10.8 App J	35.92	ensures proper decay-in- storage method
A	staff briefing for radiopharmaceutical therapy	RG 10.8 App K	35.304	ensures ancillary staff is briefed on safety measures
A	staff briefing for brachytherapy	RG 10.8 App L	35.404	ensures ancillary staff is briefed on safety measures
A	ventilation, control of gases	RG 10.8 App M	35.90, .205	avoids unnecessary worker exposure to gases
C	amendments	RG 10.8 App N	35.17	gives notice as to when amendments are needed
A	ALARA program	RG 10.8 App O	35.30	requires management commitment to ALARA