

**Veterans
Administration**

November 5, 1985

In Reply Refer To:

523-114

License No. 20-00671-02
Docket No. 030-01815John E. Glenn
Nuclear Materials Safety & Safeguards - Section B
U.S. Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, PA 19406

Dear Dr. Glenn,

The U.S. NRC has recently issued an amendment to our Materials License No. 20-00671-02. A copy of the amended licence reached my office on October 9, 1985.

Upon review of the license, two items were discovered which I believe are in significant error. Conditions 13D and 13E each set the limit of detection for sealed source leak tests at 0.0005 microcuries. I believe that this is low by an order of magnitude and that the appropriate limit of detection is 0.005 microcuries as stated in Condition 13B.

Please advise me as to the validity of the limits of detection set in Conditions 13D and 13E.

Your prompt attention to this matter is appreciated.

Sincerely,

James C. Harrington
Radiation Safety Officer (114)8602130479 851115
REG1 LIC30
20-00671-02 PDR

FREE EXEMPT

RECEIVED-REGION 1
NOV 12 1985

ML10

"OFFICIAL RECORD COPY"

NOV 12 1985