

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-432), and Title 10, Code of Federal Regulations, Chapter 1, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee

1. Bozeman Deaconess Hospital
Radiation Therapy Department

2. 22 West Beall
Bozeman, Montana 59715

3. License number 25-10994-03

4. Expiration date October 31, 1987

5. Docket or
Reference No. 030-20335/25-14896-01

6. Byproduct, source, and/or
special nuclear material:

A. Cobalt-60

7. Chemical and/or physical
form

A. Teletherapy sealed
sources (Picker Corp.
Model P3800A
or P3802A or Neutron
Products, Inc. Model
NPI-20-3000)

8. Maximum amount that licensee
may possess at any one time
under this license

A. 6,000 curies
(2 sources of
not more than
3,000 curies
each)

9. Authorized use

A. One source to be used in a Picker Corp. Model 6235 (411/60) teletherapy unit for treatment of humans. One source in its shipping container to be in possession of the licensee as necessary for the replacement of the source in the teletherapy unit only.

8512060170 850416
REG4 LIC30
25-10994-03 PDR

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-10994-03

Docket or Reference number
030-20335

CONDITIONS

10. Licensed material shall be used only in the teletherapy room at Bozeman Deaconess Hospital, 22 West Beall, Bozeman, Montana.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of Charles A. Kirkpatrick, M.D., Dail F. Lodge, M.D., or Verner Albertsen, M.D.
13. The teletherapy facility shall be provided with a system permitting continuous observation of the patient from outside the treatment room during patient irradiation.
14. A. Teletherapy sources shall be tested for leakage at intervals not to exceed 6 months. Records of test results shall be kept in units of microcuries and maintained for inspection by the Commission. In the absence of a certificate from a transferor indicating that a test has been made within 6 months prior to the transfer, a source received from another person shall not be used until tested for leakage.
B. The test shall be sufficiently sensitive to detect 0.05 microcurie of contamination on the test sample.
C. The test sample shall be taken from selected accessible surfaces of the teletherapy head. The selected accessible surfaces should be those surfaces on which one might expect contamination (if there were to be leakage) to accumulate and shall include the inner surface of the most frequently used treatment cones or beam collimating device. The test sample shall be taken with the source in the "off" position.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-10994-03

Docket or Reference number
030-20335

14. (continued)

- D. If the test reveals the presence of 0.05 microcurie or more of removable contamination, the licensee shall promptly take action to prevent spread of contamination and shall file a report within 5 days of the test with the U. S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Dr., Suite 1000, Arlington, Texas 76011 describing the test results and the corrective action taken.
15. Prior to initiation of a treatment program, each teletherapy unit shall be equipped with electrical or mechanical stops limiting use of the primary beam of radiation so as to ensure compliance with Section 20.105(b), Title 10, Part 20, Code of Federal Regulations, Chapter 1, "Standards for Protection Against Radiation," as evidenced by a radiation survey. Necessary use restrictions shall be fully described in radiation survey reports submitted in accordance with Condition No. 18.
16. A set of written emergency instructions shall be posted at the teletherapy machine control. These instructions shall inform the machine operator of the procedure to be followed should he be unable to turn the machine's primary beam of radiation "off" with the controls outside the treatment room.
17. A. Access to the teletherapy room shall be controlled by a door at each entrance. Such doors shall be normally closed.
- B. Each entrance to the teletherapy room shall be equipped with an electrical interlock system that will turn the teletherapy machine's primary beam of radiation "off" immediately upon opening of any entrance door. The interlock system shall be connected in such a manner that the teletherapy machine's primary beam of radiation cannot be turned "on" until all treatment room entrance doors are closed and the beam "on-off" control is reset at the control panel.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-10994-03

Docket or Reference number
030-20335

17. (continued)

C. Electrical interlocks on entrance doors to the teletherapy room shall be tested for proper operation at least once every 6 months. Records of test results shall be maintained for inspection by the Commission.

D. In the event of malfunction of any door interlock, the teletherapy machine control shall be locked in the "off" condition and not used, except as may be necessary for repair or replacement of the interlock system, until the interlock system is shown to be functioning properly.

18. Prior to initiation of a treatment program, and subsequent to each installation of a teletherapy source, radiation surveys and tests shall be performed in accordance with the following:

A. A radiation survey shall be made of:

- (1) The teletherapy source housing, with the teletherapy source in the "off" position. The maximum and average radiation levels at one meter from the teletherapy source in the "off" position shall not exceed 10 milliroentgens per hour and 2 milliroentgens per hour, respectively.
- (2) All areas adjacent to the treatment room with the teletherapy source in the "on" position. The survey, except Item c., shall be performed with a phantom in the primary beam of radiation and shall clearly establish:
 - a. That radiation levels in restricted areas are not likely to cause personnel exposure in excess of the limits specified in Section 20.101, Title 10, Part 20, Code of Federal Regulations, Chapter 1, "Standards for Protection Against Radiation" (10 CFR 20).

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-10994-03

Docket or Reference number
030-20335

18. (continued)

- b. That quantities of radiation in unrestricted areas do not exceed the limits specified in Section 20.105(b), 10 CFR 20.
 - c. The intensity of the primary beam of radiation at a specified distance from the teletherapy source.
8. Tests shall be made to determine proper operation of:
- (1) Electrical interlocks on entrance doors to the teletherapy treatment room.
 - (2) The teletherapy source "on-off" indicators, both at the source housing and on the teletherapy machine control panel.
 - (3) Electrical or mechanical stops installed for the purpose of limiting using of the primary beam of radiation (restriction of source housing angulation or elevation, carriage or stand travel and operation of the beam "on-off" mechanism).
 - (4) The teletherapy treatment timing device.
- C. A report of the results of the above surveys and tests shall be sent in duplicate to the U. S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Dr., Suite 1000, Arlington, Texas 76011, not later than 30 days following each installation of a teletherapy source.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-10994-03

Docket or Reference number
030-20335

19. A. Any changes made in the treatment room shielding, location of the unit within the treatment room, or use of the teletherapy unit that could result in increased radiation levels in areas outside the teletherapy treatment room shall be evaluated by a radiation survey made in accordance with Condition 18., and reported to the Commission within 30 days following completion of the change(s).
- B. Relocation of the teletherapy unit to a new facility is not permitted without prior approval of the plans and details by the Commission. Following such approval and relocation, a radiation survey shall be made in accordance with Condition 18., and reported to the Commission within 30 days after completion of the move.
20. The licensee shall comply with the requirements in Sections 35.21 through 35.27, inclusive, of Title 10, Chapter 1, Code of Federal Regulations, Part 35, "Human Uses of Byproduct Material."
21. The following shall be performed only by persons specifically authorized by the Commission or an Agreement State to perform such services:
 - A. Installation, relocation, or removal of teletherapy units containing sources.
 - B. Source exchange.
 - C. Any maintenance or repair operations on a teletherapy unit involving work on the source drawer, the shutter, or other mechanism that could expose the source, reduce the shielding around the source, or compromise the safety of the unit and result in increased radiation levels.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-10994-03

Docket or Reference number
030-20335

22. Pursuant to Title 10, Chapter 1, Code of Federal Regulations, Part 40, "Domestic Licensing of Source Material," the licensee is authorized to possess, use, transfer, and import the uranium contained as shielding material in the teletherapy units authorized by this license.
23. For a period not to exceed 60 days in any calendar year, a visiting physician is authorized to use licensed material for human use under the terms of this license, provided the visiting physician:
 - A. Has the prior written permission of the hospital's administrator and its Radiation Safety Committee, and
 - B. Is specifically named as a user on a Nuclear Regulatory Commission license authorizing human use, and
 - C. Performs only those procedures for which he is specifically authorized by a Nuclear Regulatory Commission license.

The licensee shall maintain for inspection by the Commission, copies of the written permission specified in Item A. above and of the license(s) specified in Items B. and C. above. These records shall be maintained for 5 years from the time the licensee grants its permission under Item A. above.

24. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in:

letter received November 17, 1981, signed by Thomas A. Cherewick
letter dated March 5, 1982, signed by Thomas A. Cherewick
application dated March 30, 1982, signed by Thomas A. Cherewick
letter with enclosures dated August 26, 1982, signed by Dail F. Lodge, M.D.
ALARA program dated March 30, 1982, signed by Dail F. Lodge, M.D.
letter dated February 26, 1985, signed by Dail F. Lodge, M.D.
letter dated February 20, 1985, signed by Edwin E. Dahlberg

The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date APR 16 1985

Original Signed By
C. L. Cain
By
Nuclear Materials Safety Section
Region IV
Arlington, Texas 76011

Official Record Copy

11 ML40