

SAFETY INSPECTION

Page 1 of 2

1. LICENSEE

SGS Industrial Services
2506 Davis Street
San Leandro, CA 94577

2. REGIONAL OFFICE

REGION V
U S NUCLEAR REGULATORY COMMISSION
1450 MARIA LANE SUITE 210
WALNUT CREEK CA 94596-5368

3. DOCKET NUMBER(S)

030-33390

4. LICENSE NUMBER(S)

04-29067-02

5. DATE OF INSPECTION

December 10, 1996
February 5, 1997

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

☐ 1. Within the scope of this inspection, no violations were observed.

☒ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.

☒ 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.

☐ A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b),(c),(d),(e) or 34.42.

☐ B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____

☐ C. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____

☐ D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.

☐ E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____

☒ F. On several occasions in 1996, the licensee had used a radiographic exposure device (Amersham Model 660B, S/N B1312) without a durable, legible label bearing the licensee's name, address, and telephone number, as required by 10 CFR 34.20.

150022

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE - LICENSEE

DATE

SIGNATURE - NRC INSPECTOR

DATE

[Signature]

2/5/97

[Signature]

2/5/97

SAFETY INSPECTION

Page 2 of 2

1. LICENSEE
SES Industrial Services
2506 Davis Street
San Leandro, CA

2. REGIONAL OFFICE
REGION V
U S NUCLEAR REGULATORY COMMISSION
1450 MARIA LANE SUITE 210
WALNUT CREEK CA 94596-5368

3. DOCKET NUMBER(S)

030-33290

4. LICENSE NUMBER(S)

04-29067-62

5. DATE OF INSPECTION

Dec. 10, 1996

Feb 5, 1997

3. (Continued)

G.

H.

I.

☒ 4. The violations listed below are not being cited because they were self-identified, and corrective action was or is being taken, and the remaining criteria in 10 CFR 2, App. C, to exercise discretion were satisfied.

☒ A.

The licensee failed to maintain records of the storage surveys conducted at the licensee's facility in Kapolei, Hawaii on June 27, July 9, and November 20, 1996. 10 CFR 34.43(d).

☒ B.

The licensee failed to post a high radiation area warning sign on July 17-18, 1996 during radiography in Kenai, Alaska 10 CFR 34.42 and 10 CFR 20.1902(a), (b).

☐ C.