



Borough of Pottstown

City Hall, 241 East King Street
Pottstown, Pennsylvania 19464
(215) 326-3100

November 4, 1985

Mr. John D. Kinneman, Chief
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards
United States Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, PA 19406

Re: Docket #99990001
General License Inspection Report #9999001/85-31

Dear Mr. Kinneman:

In response to your October 9 letter concerning an inspection of a nuclear measuring device located at the Pottstown Waste Treatment Plant on Moser Road in Pottstown, be advised that the borough has retained the firm of Radiation Management Corporation of Philadelphia to conduct an inspection of the above identified device in accordance with regulations.

The firm will establish whether the gauge is leaking and advise the borough accordingly if any corrective action is necessary.

In terms of corrective steps, the borough plans to either inspect this gauge at three-year intervals from the date of the above identified inspection if it does not decide to dispose of the device in the meantime. The device has not been used since its installation and is of no value in its initial purpose of measuring sludge viscosity.

We expect to achieve full compliance in terms of inspection by November 8, 1985.

Very truly yours,

Thomas J. Harwood, Jr.
Borough Manager

8512060258 851129
IE GA999 ELOPOTT
99990001 PDR

CC: Mr. John Jones, Chief Waste Treatment Plant Operator
Mr. Wayne Vogel, Radiation Management Incorporated