

CORRECTED COPY

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee			
1. Island Creek Coal Company Box 840		3. License number	47-15019-02
2. Holden, West Virginia 25625		4. Expiration date	December 31, 1990
		5. Docket or Reference No.	030-28969
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license	
A. Cesium 137	A. Sealed sources (Troxler Electronics Dwg. No. A-102112)	A. No single source to exceed 10 millicuries	
B. Americium 241	B. Sealed neutron sources (Troxler Electronics Dwg. No. A-102451)	B. No single source to exceed 50 millicuries	
9. Authorized use			
A. and B. For use in Troxler Electronics Models 3401 or 3411 portable moisture/density gauges to measure properties of material.			

CONDITIONS

10. Licensed material may be stored at the licensee's facilities on State Route 119, Holden, West Virginia, and may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. Licensed material shall be used by, or under the supervision and in the physical presence of, Danny Cox or individuals who have been trained as specified in application dated October 15, 1985. The licensee shall maintain records of individuals designated as users.

8602180094 860129
REG2 LIC30
47-15019-02 PDR

11 ML20

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

47-15019-02

Docket or Reference number

030-28969

(cont'd)

CONDITIONS

12. A.(1) The source(s) specified in Items 7.A and B shall be tested for leakage and/or contamination at intervals not to exceed 6 months. Any source received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- B. Any source in storage and not being used need not be tested. When the source is removed from storage for use or transfer to another person, it shall be tested before use or transfer.
- C. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the source shall be removed from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the date the leak test result is known with the U. S. Nuclear Regulatory Commission, Region II, Division of Radiation Safety and Safeguards, Nuclear Material Safety Section, 101 Marietta Street, Suite 2900, Atlanta, Georgia 30323. The report shall specify the source involved, the test results, and corrective action taken. Records of leak test results shall be kept in units of microcuries and shall be maintained for inspection by the Commission. Records may be disposed of following Commission inspection.
- D. Tests for leakage and/or contamination shall be performed by the licensee or by other person specifically licensed by the Commission or an Agreement State to perform such services.
13. The licensee may transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
14. Sealed sources containing licensed material shall not be opened.
15. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintain for 2 years from the date of each inventory.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number 47-15019-02

Docket or Reference number 030-28969

(cont'd)

CONDITIONS

16. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- ° Application dated October 15, 1985.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date JAN 29 1986By EARL G. WRIGHT
Earl G. Wright
Region II, Nuclear Materials
Safety Section
101 Marietta Street, Suite 2900
Atlanta, GA 30323