

*File*

SEP 25 1984

Northern Medical Imaging, Inc.  
ATTN: Dennis Clemenson  
President  
P.O. Box 1479  
Bemidji, MN 56601

Gentlemen:

We are in receipt of your letter dated July 16, 1984 requesting an amendment to NRC License No. 22-16328-01. Please supplement the information contained in the letter by responding to the following:

1. Describe the type of building in which your facilities in Brainerd, Minnesota will be located, e.g. office building, warehouse, etc.  
  
If the building is a house in a residential area, please confirm the following:
  - a. The building will not be occupied as a residential dwelling.
  - b. The possession of radioactive material in the house will not conflict with local codes or zoning laws.
2. Describe how radioactive materials in the building will be secured from unauthorized use or removal. How will access to the building be controlled?
3. Describe the type of shielding that will be used for:
  - a. Use and storage of Tc-99 generators.
  - b. Storage of radiopharmaceuticals.
  - c. Storage of radioactive waste.
  - d. Preparation and dispensing of Groups III kit radiopharmaceuticals.
4. Specify the instrumentation that will be available for use at your new facilities, e.g. survey meters, dose calibrator, etc.

We will continue the review of your request upon receipt of this information. Please reply in duplicate within 30 days and reference Control No. 77140.

Sincerely,

8512060297 841028  
REG3 LIC30  
22-16328-01 PDR

Original Signed By  
B. J. Holt  
Materials Licensing Section

RIII

Holt/cm *BH*  
09/19/84 *9/20/84*