

SAFETY INSPECTION

Page 1 of 1

1. LICENSEE

VETERAN AFFAIRS MEDICAL CENTER
2300 RAMSEY ST.
FAYETTEVILLE, NC 28301

2. REGIONAL OFFICE

REGION II
U S NUCLEAR REGULATORY COMMISSION
101 MARIETTA STREET NW SUITE 2900
ATLANTA GA 30323

3. DOCKET NUMBER(S)

030-02634

4. LICENSE NUMBER(S)

32-13654-01

5. DATE OF INSPECTION

02/25/97

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☐ 1. Within the scope of this inspection, no violations were observed.
- ☐ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- ☒ 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.
- ☐ A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b),(c),(d),(e) or 34.42.
- ☐ B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____.
- ☐ C. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____.
- ☐ D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
- ☐ E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____.
- ☒ F. LK 16 REFERENCES DOCUMENTS TIED TO THE LICENSE, ITEM 10.4 OF APPLICATION DATED FEBRUARY 25, 1991 REQUIRES THE IMPLEMENTATION OF REG. GUIDE 10.8, APPENDIX I. ITEM 6 OF APPENDIX I STATES THAT FOOD, DRINK OR PERSONAL EFFECTS NOT BE STORED IN AREAS OF RADIOACTIVE MATERIAL USE OR STORAGE. CONTRARY TO THIS, A CAN OF ORANGE JUICE WAS FOUND IN THE HOT LAB REFRIGERATOR DURING THE INSPECTION.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE - LICENSEE <i>AMM</i>	DATE 2/25/97	SIGNATURE - NRC INSPECTOR <i>B. A. Palm</i>	DATE 02/25/97
------------------------------------	-----------------	--	------------------