

NOV 7 1985

Docket No. : 030-09518  
License No.: 46-10100-02  
Control No.: 70279

Environmental Protection Agency  
Region 10 Laboratory  
P.O. Box 549  
Manchester, Washington 98353

Attention: Mr. Michael M. Johnston, Chief  
EPA Region 10 Laboratory

Gentlemen:

Enclosed is the NRC license amendment which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

You should note that Hewlett Packard Model Numbers were used to describe the New England Nuclear detector cells which are used in the Model 5800 series gas chromatographs.

Also, the only titanium tritide foil registered under Tracor's name is a Model 111020 cell with a 150 millicurie limit. Therefore, your license was amended to allow the possession and use of these cells.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely, ,

8602180403 851107  
REG5 LIC30  
46-10100-02 PDI

Beth A. Riedlinger  
Health Physicist (Licensing)  
Nuclear Materials Safety Section

Enclosure: Attachment A  
Amendment No. 7 to License No. 46-10100-02

OFFICE	RV						
NAME	Riedlinger:fr						
DATE	11/1/85						