

SEP 19 1985

Dielman Consultants, Inc.
Box 10305
Chicago, IL 60610

Gentlemen:

Enclosed is Amendment No. 03 to your NRC License No. 12-17162-02 in accordance with your request.

You are reminded that the room in which you will store your calibration device must comply with the requirements of 10 CFR 20.203(e) and should be labeled accordingly and must meet the definition of a restricted area as described in 10 CFR 20.3(14).

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
William J. Adam, Ph.D.
Materials Licensing Section

Enclosure(s): Amendment No. 03

8510030406 850919
REG3 LIC30
12-17162-02 PDR

RIII
wja
Adam/cm
09/18/85