

MATERIALS LICENSE

Amendment No. 41

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. V.A. Edward Hines, Jr., Medical Center	In accordance with application dated June 15, 1984
2. Hines, IL 60141	3. License number 12-01087-07 is amended in its entirety to read as follows:
	4. Expiration date September 30, 1990
	5. Docket or Reference No. 030-01391
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form
A. Any byproduct material listed in Groups I and II of Schedule A, Section 35.100 of 10 CFR 35	A. Any radiopharmaceutical listed in Groups I and II of Schedule A, Section 35.100 of 10 CFR 35
B. Any byproduct material listed in Group III of Schedule A, Section 35.100 of 10 CFR 35	B. Any form listed in Group III of Schedule A, Section 35.100 of 10 CFR 35
C. Any byproduct material listed in Group IV of Schedule A, Section 35.100 of 10 CFR 35	C. Any radiopharmaceutical listed in Group IV of Schedule A, Section 35.100 of 10 CFR 35
D. Any byproduct material listed in Group V of Schedule A, Section 35.100 of 10 CFR 35	D. Any radiopharmaceutical listed in Group V of Schedule A, Section 35.100 of 10 CFR 35
E. Any byproduct material listed in Group VI of Schedule A, Section 35.100 of 10 CFR 35	E. Any sealed source listed in Group VI of Schedule A, Section 35.100 of 10 CFR 35
	8. Maximum amount that licensee may possess at any one time under this license
	A. As necessary for uses authorized in Subitem 9.A
	B. 10 curies of each byproduct material authorized in Subitem 6.B
	C. As necessary for uses authorized in Subitem 9.C
	D. As necessary for uses authorized in Subitem 9.D
	E. 2.1 curies total for all sources authorized in Subitem 6.E

8510040028 850924
REQ LIC30
12-01087-07 PDR1/1 COPY 2
nu30

B3 1124/85

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

12-01087-07

Docket or Reference number

030-01391

Amendment No. 41

6. Byproduct, source,
and/or special nuclear
material

7. Chemical and/or
physical form

8. Maximum amount that
licensee may possess
at any one time
under this license

F. Xenon-133

F. Gas or gas in solution
that is the subject
of an active (i.e., not
withdrawn or terminated)
"New Drug Application"
(NDA) approved by FDA or
an active (i.e., not
withdrawn, terminated
or on "clinical hold")
"Notice of Claimed
Investigational Exemption
for a New Drug" (IND)
that has been accepted
by FDA

F. 2 curies

G. Any byproduct material
between Atomic Nos. 1
and 83, inclusive

G. Any

G. 50 millicuries
of each byproduct
material, except
as noted below:

Hydrogen-3	750 millicuries
Carbon-14	750 millicuries
Cobalt-60	100 millicuries
Iodine-125	750 millicuries

9. Authorized Use

A. Any diagnostic procedure listed in Groups I and II of Schedule A, Section 35.100 of Title 10, Code of Federal Regulations.

B. Preparation and use of radiopharmaceuticals for any diagnostic procedure listed in Group III of Schedule A, Section 35.100 of Title 10, Code of Federal Regulations.

C. Any therapeutic procedure listed in Group IV of Schedule A, Section 35.100 of Title 10, Code of Federal Regulations.

D. Any therapeutic procedure listed in Group V of Schedule A, Section 35.100 of Title 10, Code of Federal Regulations.

E. Any procedure listed in Group VI of Schedule A, Section 35.100 of Title 10, Code of Federal Regulations.

F. Blood flow studies. Pulmonary function studies.

G. To be used in medical research and development as defined in 10 CFR Part 30, 30.4(q).

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

12-01087-07

Docket or Reference number

030-01391

Amendment No. 41

CONDITIONS

10. Licensed material shall be used only at the licensee's facilities located at Edward Hines Jr. Medical Center, Hines, Illinois.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12.
 - A. Licensed material shall be used by, or under the supervision of, individuals designated by the Medical Radioisotope Committee, Ervin Kaplan, M.D., Chairman.
 - B. The use of licensed material in or on humans shall be by a physician as defined in 10 CFR 35.3(b).
 - C. The Radiation Protection Officer for the activities authorized by this license is Lawrence F. Case.
13.
 - A.
 - (1) Each sealed source acquired from another person and containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for contamination and/or leakage prior to use. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
 - (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
 - (3) Except for alpha sources, the periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within six months prior to the date of use or transfer.
 - B. Each sealed source fabricated by the licensee shall be inspected and tested for construction defects, leakage, and contamination prior to use or transfer as a sealed source. If the inspection or test reveals any construction defects or 0.005 microcurie or greater of contamination, the source shall not be used or transferred as a sealed source until it has been repaired, decontaminated and retested.
 - C. Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months except that each source designed for the purpose of emitting alpha particles shall be tested at intervals not to exceed three months.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

12-01087-07

Docket or Reference number

030-01391

Amendment No. 41

- D. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently or semipermanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- E. If the test required by Subsection A. or C. of this condition reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.
14. Sealed sources containing licensed material shall not be opened.
15. In lieu of using the conventional radiation caution colors (magenta or purple on yellow background) as provided in Section 20.203(a)(1), Title 10, Code of Federal Regulations, Part 20, the licensee is hereby authorized to label detector cells and cell baths, containing licensed material and used in gas chromatography devices, with conspicuously etched or stamped radiation caution symbols without a color requirement.
16. Experimental animals administered licensed materials or their products shall not be used for human consumption.
17. Patients containing Iodine 131 for the treatment of thyroid carcinoma (or patients containing therapeutic quantities of Gold 198) shall remain hospitalized until the residual activity is 30 millicuries or less.
18. The licensee is authorized to hold radioactive material with a physical half-life of less than 65 days for decay-in-storage before disposal in ordinary trash provided:
- A. Radioactive waste to be disposed of in this manner shall be held for decay a minimum of ten (10) half-lives.
- B. Prior to disposal as normal waste, radioactive waste shall be monitored to determine that its radioactivity cannot be distinguished from background with typical low-level laboratory survey instruments. All radiation labels will be removed or obliterated.
- C. Generator columns shall be segregated so that they may be monitored separately to ensure decay to background levels prior to disposal.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

12-01087-07

Docket or Reference number

030-01391

Amendment No. 41

19. The licensee may use the Lineator device for doing linearity tests of his dose calibrator provided he follows the procedures in the Atomic Products Corporation Lineator Instructions Manual dated June 20, 1983.
20. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated June 15, 1984 and letter dated September 10, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



For the U.S. Nuclear Regulatory Commission

Date

SEP 24 1985

Original Signed
By William J. Adam
Materials Licensing Section, Region III