

Mobil Pipe Line Company

October 17, 1985

Nuclear Regulatory Commission
Attn: Mrs. M. Taylor

Docket #9990001

Instrument wipe test was performed on 6-7-85. The radiation level was found to be less than .005 uci.


New source was installed, radiation survey and shutter mechanism inspection by Princeton Gamma-Tech on 7-30-85.

Wipe test performed on 8-2-85, radiation level found to be less than .005 uci.

Company personnel have placed wipe test due date sticker on instrument and have been instructed on how to perform the test to insure compliance with NRC policy.

Full compliance has been achieved as of 8/85.

Sincerely,



R. E. Sheaffer
Station supervisor

8512100184 851125
IE QA999 EMVMOPIP
99990001 PDR

12 3 41 42 130 500
RECEIVED-REGON 1
1985 OCT 24 PM 3:21