



# RESEARCH DYNAMICS INCORPORATED

637 REDNA TERRACE, UNIT 1, CINCINNATI, OHIO 45215. (513) 772-8400



December 5, 1985

RDI-NRCO-005

Mr. Gary G. Zech, Chief  
Vendor Program Branch  
Division of Quality Assurance  
Vendor and Technical Training Center Programs  
Office of Inspection and Enforcement  
United States Nuclear Regulatory Commission  
Washington, D.C. 20555

Re: Docket No. 99900930/85-01

Dear Mr. Zech:

This is in response to the Notice of nonconformance, Appendix A, Docket No. 99900930/85-01, dated November 13, 1985.

Attached as Appendix A is our Statement in response to the notice of nonconformance.

We feel that the corrective actions to be initiated ensure conformance with Criterion V of Appendix B to 10CFR Part 50.

We trust that this is a satisfactory response to Docket No. 99900930/85-01; however, if you need further information, we will be happy to comply.

Yours truly,

RESEARCH DYNAMICS INCORPORATED

*F. F. Suagee Jwa*

Dr. Freeman F. Suagee,  
Vice President and General Manager

FFS/ja

Encl. Appendix A

cc: Dr. J.N. Anno  
J.W. Anno (Document Controller)  
Dr. M.J. Pool  
L.A. Rieser  
D.A. Tierrey

8512100220 851205  
PDR GA799 EECRESY  
99900930 PDR

*IE09*  
*11*

APPENDIX A

Statement in Response to Notice of Nonconformance, Docket No.  
99900930/85-01, Research Dynamics Incorporated, Cincinnati, Ohio

With respect to the item listed under the category of "Notice of Nonconformance, " the following corrective actions will be taken:

1. The RDI Procedure for Calibration of Test Pressure Transmitters, RDI-INT-060, will be revised to contain an acceptance criterion of plus or minus 2% of full scale reading.
2. A review and approval of all calibration data by the Quality Assurance Officer will be required.

The above corrective actions will be completed by January 1, 1986.