



MENASHA CORPORATION

August 8, 1985

USNRC Region III
Materials Licensing Section
799 Roosevelt Road
Glen Ellyen, Illinois 60137

Gentlemen:

The purpose of this letter is to request two ammendments to our Materials License #21-14922-01.

First, one of the five Model LS-101 source holders and its CS-137 source was removed from service and put into storage. The holder has been locked in the closed position. This unit had been mounted on a rail and used to determine level in the digester tube. The LS-101 source holder was replaced with source which is held under the general license of Nuclear Research Corporation.

We also ask that Sytec Corporation be added as a service company permitted to perform leak tests on our sources.

Please call if there are any questions.

Sincerely,

Menasha Corporation
Otsego Paperboard Division

Mark T. Reed

Mark T. Reed
Technical Manager

kj

8510040324 850808
REG3 LIC30
21-14922-01 PDR

RECEIVED

AUG 12 1985

REGION III

Applicant	<i>Aug 21</i>
Check No.	<i>4881932</i>
Amount/Fee	<i>400</i>
Type of Fee	<i>31 and</i>
Date Check Rec'd	<i>8/21/85</i>
Received By	<i>[Signature]</i>

U.S. N.R.C.
LIC. FEE MGMT. BRANCH

85 AUG 21 10:45

RECEIVED

AUG 12 1985

ML-4
111

OTSEGO PAPERBOARD MILL FARMER STREET • OTSEGO, MICHIGAN 49078 • PHONE 616-692-6141

CONTROL NO. 7-9554 *401 del*

NOTE TO: License Fee Management Branch, ADM

FROM: Region III

SUBJECT: VOIDED APPLICATION

Control Number 79554

Applicant Menasha Corp.

Date Voided 9/5/85

Reason for Void Licensee is already authorized
for request - Amendment not
necessary - licensee notified of same.

Signature

J. L. Mad

Attachment:
Application

OKLFMB

no refund due -
review was initiated
to determine that by the
is not authorized to perform
leak tests (see Madena's
9/5/85 telecon record)

to Jackson
9/13/85

logged copy

ML30
1/1

CONVERSATION RECORD

TIME

9:00 @

DATE

9/5/85

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☒ OUTGOING

ROUTING

NAME/SYMBOL INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

ORGANIZATION (Office, dept., bureau, etc.)

TELEPHONE NO.

Mark T. Reed, Tech. Menasha Corp.

616/692-6141

SUBJECT

Manga.

Void Amendment request.

SUMMARY

① Info. provided is fine but not necessary on an amendment. E/c. 13. D. states that tests for leakage and/or contamination shall be performed by the licensee or by persons specifically licensed to do so by the Com. or an agreement state. - Sytec Corp. is not authorized to perform leak tests on your gauges or for you as a licensee. - You should make sure the co. which provides this service is authorized and gives you that authorization in writing for future inspections. (Service licensee).

② It is not necessary to provide us info. on transfers of licensed material in the form of an amend. Keep all paperwork in file for future insp. etc.

ACTION REQUIRED

Void amendment request # 79554

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

J.K. Made

9/5/85

ACTION TAKEN

SIGNATURE

TITLE

DATE