

RADIATION PROTECTION SERVICES, INC.

P.O. Box 2359, Darien CT. 06820

October 8, 1985

MS 16
K2

Mr. Jack Davis
Nuclear Materials Safety Section A
Division of Radiation Safety & Safeguards
U.S. Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, PA 19406

Re: License Application - RPS, Inc.
Docket No. 030-22112
Mail Control No. 03454

Dear Mr. Davis:


We are in receipt of your letter dated 30 September 1985 in which additional information is required to continue review of our license application. The requested information is as follows:

(1) We, hereby, confirm that we will only provide calibrations of low and medium range survey meters.

(2) We, hereby, confirm that if a scale is not or cannot be calibrated without repair, a calibration certificate will not be issued.

If any additional information is required, please do not hesitate to contact my office.

With best regards,


Stuart R. Korchin, PE
President

SRK/kjc

8512100308 851029
REG1 LIC30
06-20811-01 PDR

03454

"OFFICIAL RECORD COPY"

ML10

OCT 11 1985