

SEP 30 1985

Docket No. 030-22112
Control No. 03454

Radiation Protection Service, Inc.
ATTN: Mr. Stuart R. Korchin, PE
President
P.O. Box 2359
Darien, Connecticut 06820

Gentlemen:

This is in reference to your request in a letter dated September 9, 1985 for a byproduct material license. In order to continue our review, we need the following additional information:

1. In item 1. of your letter, you provide data that indicates that you will not be able to calibrate high range survey meters in accordance with the guidelines contained in the American National Standard for Radiation Protection Instrument Calibration (ANSI N323-1978). Please confirm that you will provide only calibrations of low and medium range survey meters.
2. In item 5. of your Survey Meter Calibration Procedures, you state if a scale is not calibrated, a note will be attached to the calibration certificate. Please confirm that if a scale is not or cannot be calibrated without repair, a calibration certificate will not be issued.


We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 03454.

Sincerely,

Original Signed By:
Jack Davis

Jack Davis
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

RI:DRSS
Davis/rw
9/30/85


RI:DRSS
Kinneman
9/30/85

OFFICIAL RECORD COPY

ML103 DAVIS 9/23/85 - 0001.0.0
09/26/85

8512100310 851029
REG1 LIC30
06-20811-01 PDR

ML103