



14650 Southlawn Lane, Suite 4
Rockville, Maryland 20850
(301) 424-3090

September 9, 1985

MS 16
KO

United States Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, Pennsylvania 19406

Attention: Mr. John D. Kinneman

Re: Mail Control No. 103959
ITL, Inc. Byproduct Material License Application

Dear Mr. Kinneman:

In response to your letter of August 23, 1985 requesting additional information relative to our firm's application dated June 10, 1985; we hereby offer the following additional information:

- 1) Leak testing will be performed by Robert E. Tolley of our firm. the information relative to the test kit is:

Kit Model No.: TK-3880

Name: Troxler Electronic Laboratories, Inc.

Address: Triangle Park, North Carolina

- 2) We hereby confirm that no maintenance will be performed that involves removing the source or placing it in an unshielded position.
- 3) We hereby confirm that the packaging and transport of the device will be carried out in accordance with applicable DOT regulations.
- 4) C_s137 is the correct listing in Item 8, number 2 of our application.

We trust the above information satisfactorily provides the requested additional information.

Should you have any questions regarding the above, please do not hesitate to contact our office.

Very truly yours,

ITL, Inc.

Robert E. Tolley
Robert E. Tolley
Vice President

B510300205 B50924
REG1 LIC30
19-20864-01 PDR

RET:jm
cc: (2) U.S. Nuclear Regulatory Commission

03959
"OFFICIAL RECORD COPY"

ML10

SEP 12 1985