

MERCY CATHOLIC MEDICAL CENTER
OF SOUTHEASTERN PENNSYLVANIA

October 14, 1985

Mr. Jack Davis
U.S. Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, Pa. 19406

Dear Mr. Davis:

I would like to request the following amendments to our license
37-00993-04 - Reference Docket 030-00460, as follows:

- a) Removal of teletherapy source model AMS 3802
Serial # AMS 2364 and rotational therapy unit
from Misericordia Hospital, Philadelphia, Pa.
- b) Removal of teletherapy source model Picker P 3802 M
Serial # PX 2262 and VM 60 Therapy Unit from the
Fitzgerald-Mercy Hospital, Darby, Pa.

Both sources to be removed by:

Clyde Peabody
X-Ray Equipment Company
Fort Worth, Texas

(817) 429-5099

Texas License is: TX-5-1485

- c) Possession of 11000 curies (2 sources of not more than
5500 curies each). Cobalt sealed source CEA Model COT-20.
The source is to be installed in a model number Alcon II
Rotational Unit at Misericordia Hospital, Philadelphia, Pa.

RECEIVED
85 OCT 23 AM 8:58
U.S. N.R.C.
LIC. FEE MGMT. BRANCH

Commissariat a
l'Energie
Atomique

Applicant	Oct 13
Check No.	31029
Amount/Fee Category	\$230/74
Type of Fee	AMB
Date Check Rec'd	10/28/85
Received By	J. Schenck

8512180478 851118
REG1 LIC30
37-00993-04 PDR

85 OCT 16 100 5001

"OFFICIAL RECORD COPY"

ML10

104515

MISERICORDIA DIVISION
54th Street and Cedar Avenue
Philadelphia, Pa. 19143
(215) 748-9000

OCT 17 1985

The installation will be done by:

CGR Medical Corporation
10150 Old Columbia Road
Columbia, Maryland 21046


Maryland License # MD-27-028-01

The Cobalt room at Misericordia Hospital was designed to use up to 8000 curies of Cobalt. No change in shielding is expected.

The Alcon II has a retractable primary shield. We have requested that this shield be locked in the out position.

Our anticipated schedule is hopefully December 16, 17, 1985 - removal of Cobalt Units. December 23, 1985 - installation of new units.

Sincerely yours,


C. Jules Rominger, M.D.

CJR/me

c: James Bierly, Ph.D.
Philip Turner
CGR Medical Corporation
Cleveland Mair
File

BETWEEN: William O. Miller, Chief
License Fee Management Branch
Office of Administration

John E. Glenn, Chief
Nuclear Materials Section B
Division of Engineering and
Technical Programs

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: Mercy Catholic Medical Center

Application Dated: 10/14/85

Control No.: 104515

License No.: 37-00993-04

2. FEE ATTACHED

Amount: 0 / \$230

Check No.: 0 / 31029

3. COMMENTS

02300

8/90

Signed Brenda Pilatchek

Date 10/17/85

B. LICENSE FEE MANAGEMENT BRANCH

1. Fee Category and Amount: 7A (\$230)

2. Correct Fee Paid. Application may be processed for:

Amendment ☒

Renewal ☐

License ☐

Signed B Jackson

Date 11/1/85

Verbal OK
given earlier
(Rush)

~~Fee Needed~~