



TOBEY HOSPITAL, INC.
43 HIGH STREET
TELEPHONE (617) 295-0880
WAREHAM, MASSACHUSETTS 02571

RECEIVED

85 NOV -8 1

U.S. NRC
12. FILE MMT. BRANK

October 30, 1985

Dr. John Glenn, Chief
United States Nuclear Regulatory
Commission
Nuclear Materials Safety Section B
631 Park Avenue
King of Prussia, Pennsylvania 19406

Dear Dr. Glenn:

This letter is to request amendment of NRC License #20-16392-01 issued to the Tobey Hospital, Inc. Wareham, Massachusetts for the following:

1. Add Bruce Leiter, M.D. as an individual user to the License for Groups I, II and III. Attached is a copy of Condition 12 of the Boston City Hospital License on which Dr. Leiter was an individual user.
2. Use of the Calicheck device for performing linearity tests of the dose calibrator in accordance with the procedures in the Calcorp, Inc., Manual dated March 2, 1982.
3. Remove James J. Condon, M.D. as an individual user on the License. He is no longer affiliated with Tobey Hospital, Inc.
4. Change calibration of survey instruments to the Nuclear Instrument Co., 65 Grove Street, Rockland, Massachusetts, NRC License #20-16972-01.
5. Use of Seimens Gammasonics, Inc., 2000 Nuclear Drive, Des Plaines, Illinois, 60018, NRC License #12-00369-01 to perform counting of leak tests on the sealed sources.

If you require any additional information, it will be sent to you immediately.

Sincerely,

Charles S. Kinney
Charles S. Kinney
President

Enclosure

CSK:pmk

8512190027 851118
REG1 LIC30
20-16392-01 PDR

Nov 8 - 1

Applicant	
Check No.	11884
Amount/Fee	\$120.70
Type of Fee	Amtd
Date Check Rec'd	11/8/85
Received By	Jacques

85 NOV 11 - 11:00 AM '85

RECEIVED-REGISTRATION

"OFFICIAL RECORD COPY"

104591

ML10

NOV 01 1985