

OCT 25 1985

Docket No. 30-19307

License No. 37-13360-03

Analytical Instrument Development, Inc.
ATTN: Mr. E. Neel
President
Rt. 41 and Newark Road
Avondale, Pennsylvania 19311

Gentlemen:

Subject: Inspection No. 85-01

This refers to your letter dated July 18, 1985, in response to our letter dated July 10, 1985.

Thank you for informing us of the corrective and preventive actions you documented in response to our correspondence. However, your letter has been reviewed and additional information is found to be needed.

With respect to your response to item 2 of the Notice of Violation, you are advised that License Condition 11 requires that, whenever one of your instruments is taken into the field, you or your Radiation Safety Officer (Mr. Debbrecht) be physically present. This violation refers to our finding that your sales representative (Mr. J. McAfee) transported a demonstration instrument from your place of business to a trade show, and neither you or Mr. Debbrecht were physically present at that location. You may request an amendment to your license to permit him or others to use an instrument in the field but you are required to comply with Condition 11 of your license until such an amendment has been approved. In your reply to this letter, please confirm that licensed materials will only be used at temporary job sites while in the physical presence of yourself or Frederick Debbrecht.

Your letter states that you are no longer using the proportional gas flow beta counter to count wipes; instead you are using the services of a consultant. License Condition 14 requires that you maintain an operable proportional gas flow beta counter. If you wish to change your procedure for contamination monitoring, you must request an amendment to your License to reflect this change. Requests for license amendments must be submitted, with the appropriate amendment fee, to the USNRC, Nuclear Materials Safety and Safeguards Branch, Region I, King of Prussia, Pennsylvania 19406.

We regret the omission of the results of the smear surveys taken by the inspector. No removable Hydrogen-3 contamination was detected in your unrestricted areas.

8510310078 851025
REG1 LIC30
37-13360-03 PDR

OFFICIAL RECORD COPY

RL ANALYTICAL - 0001.0.0
10/15/85

RETURN ORIGINAL TO
REGION I

1E:07

OCT 25 1985

2

Your letter asked for clarification regarding our letter's reference to the Paperwork Reduction Act of 1980. This Act is designed to prevent Federal Regulatory Agencies from requesting unnecessary paper work from the public. The purpose of this paragraph in our letter is to state that the response required of you is not subject to, and, therefore, is in conformance with the requirements of that Act. The request contained in this letter is also not subject to that Act.

Please submit to this office within fourteen (14) days of receipt of this letter a written statement containing the requested information.

Your cooperation with us in this matter is appreciated.

Sincerely,


Original Signed By:
John D. Kinneman

John D. Kinneman, Chief
Nuclear Materials Safety Section A,
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
Commonwealth of Pennsylvania

bcc:
Region I Docket Room (w/concurrences)

RI:DRSS
Ladun/ca
10/16/85


RI:DRSS
Kinneman
10/14/85

OFFICIAL RECORD COPY

RL ANALYTICAL - 0002.0.0
10/15/85