

*Nuclear + Radiologic
Imaging Phys.*
SUBHASH C. KHULLAR, M.D., P.C.

Troy Professional Building
215i Livernois
Troy, MI 48083
362-1360

DMB COPY

30-18655

21-24470-01

S.C. Khullar, M.D.
F.C. Stebner, M.D.

December 4, 1985

D. G. Wiedeman, Chief
Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Dear Mr. Wiedeman:

In response to your letter dated November 8, 1985, regarding the inspection recently conducted at our Livernois office and the violations referred to in your letter, as a result of this inspection.

Several areas of non-compliance were found as outlined in your letter. Corrective action has been taken as follows:

1. The Bicron 2000 survey meter has been returned to the manufacturer for recalibration, specifically the "X-100 scale". As the initial calibration was performed by the manufacturer and as the instrument is still under warranty, Bicron will recalibrate the instrument and re-issue a new calibration certificate, or a chart, graph, or response factor will be prepared as required by Appendix D.
2. The off-hours delivery of radioactive material to the lobby of our Troy office has been discontinued. The supplier and technician have been notified that delivery will be made during normal business hours, with the presence of the technician on site to receive the delivery and store the delivery in the posted area of the facility. The delivery area in the lobby was removed the day of the inspection and surveyed, and no further deliveries will therefore be left in the lobby area.
3. The radioactive waste canister in the imaging room was labeled the same day as the inspection. An identifying label is therefore on the container and will remain so.

Therefore, items numbered two and three were corrected the same day as the inspection. We expect to receive our survey meter from the manufacturer, recalibrated, within the next week. If there are any further questions or information which you would require, please feel free to contact me. Thank you very much.

Sincerely,



F. Carlyle Stebner, M.D.
Deputy Radiation Safety Officer

FCS/vn

8512190198 851204
REG3 LIC30
21-24472-01 PDR

IE07
110
DEC 9 1985