

DMB COPY

DEC 16 1985

Marion Memorial Hospital
ATTN: Merle Aukamp
Administrator
917 West Main Street
Marion, IL 62959

License No. 12-17600-01

Gentlemen:

Thank you for your letter dated November 29, 1985, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated November 8, 1985.

We have considered the matter in item 2 of your letter, and conclude that a survey of the overpack containing the molybdenum/technetium generator and the xenon-133 gas meets the intent of the survey requirement. As a result of this additional information we are withdrawing item 2 from our Notice of Violation, dated November 8, 1985.

From the telephone conversation with George Murphy, M.D. on December 6, 1985, it is our understanding the your corrective actions will include the following:

1. To prevent unauthorized removal and unauthorized storage, the xenon-133 gas trap will be stored in the camera room.
2. A quarterly review of the nuclear medicine department will be performed by the consultant physicist to ensure continued compliance.

If your understanding differs from the above please contact this office.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

D. J. Sreniawski

D. J. Sreniawski, Chief
Nuclear Materials Safety
Section 2

8512190211 851216
REG3 LIC30
12-17600-01 PDR

cc w/ltr dtd 11/29/85:
DCS/RSB (RIDS)

yes
RIII
OP for
Reichold/jp
12/11/85

yes
RIII
DJS
Sreniawski
12/12/85

IE07

11