

SEP 8 1982

Docket No. 030-08449
Control No. 00797

License No. 37-15029-01

Broad Street Hospital
ATTN: Miss Sonia Sokoloff
Administrator
739 South Broad Street
Philadelphia, Pennsylvania 19147

Dear Miss Sokoloff:

This is in reference to your application dated August 4, 1982, to renew License No. 37-15029-01. In order to continue our review, we need the following additional information:

You should have available in your department a high-level survey meter capable of reading up to 1 Roentgen per hour in order to measure radiation dose rates that may exist in the vicinity of Tc-99m generators, etc. Please indicate the manufacturer's name, model number, and highest level range of the instrument in your facility that will fulfill this need.

However, if you are using individual patient doses received from a radio-pharmaceutical supplier, you do not need the high-level meter. If this is the case, please so indicate and confirm that you will procure a high-level meter in the event you switch to a generator for your technetium-99m needs.

We will continue our review of your application upon receipt of this information. Please reply in duplicate and refer to Control No. 00797.

Sincerely,

Original Signed By:
Phillip C. Jerman

Phillip C. Jerman
Materials Program Section No. 2
Division of Engineering and
Technical Programs

8510310499 850930
REG. LIC30
37-15029-01
PDR

OFFICE	RI:DETP	RI:DETP			
SURNAME	Jerman/dmg	Glenn			
DATE	8/31/82	9/3/82			