

November 15, 1985

U.S. Nuclear Regulatory Commission  
Nuclear Materials Safety Section B  
Region 1  
631 Park Avenue  
King of Prussia, Pa. 19406

Subject: Inspection No. 030-17742/85-01


Gentlemen,

With regards to the letter received concerning the safety inspection by M.A. Varela, the following steps have been taken to rectify the violations.

- A) The storage shed housing the nuclear spectrometer is now locked any time the sorter leaves the immediate area. The key for the lock is held by the sorter and only authorized users of this machine.
- B) Since the issue of our license, the supervisors, Chester Moore and Ellis Jones have retired. The person using the Kevex on the day in question will be listed as an authorized user of the machine. We will have our license amended to include William McManus - (Sorter) and William C. Fisher - (Plant Manager) as authorized users and/or supervisors.
- C) The above named Ellis Jones was the control supervisor of all records pertaining to the Kevex records. Since his retirement, Shirley Lippai will be designated to keep all leak down tests records as well as personal badge records for the individuals listed above who use the machine. These records are now kept in a cabinet available for immediate inspection.

I hope that the above steps will correct the violations listed and bring us into immediate compliance.

Sincerely,



William C. Fisher  
Plant Manager

WCF/cc

8512200241 851217  
REG1 LIC30  
37-19469-01 PDR

