

JUN 19 1985

License No. 06-01060-01
Docket No. 030-01247
Control No. 103859

Bridgeport Hospital
ATTN: Michael E. Schrader
President
267 Grant Street
Bridgeport, Connecticut 06602

Gentlemen:

This is in reference to your request in a letter dated May 20, 1985 to amend License No. 06-01060-01. In order to continue our review, we need the following additional information:

1. Please identify the ventilation rate of your xenon 133 storage area and confirm that this area is under negative pressure.
2. In your letter dated May 20, 1985 you indicate that the old hot lab is being eliminated and that one of your scan rooms will also be moved. Please verify whether these areas will be released for unrestricted use. Prior to releasing these areas for unrestricted use, it will be necessary for you to perform a closeout survey (guide enclosed) of these areas and submit the results for our review. In submitting the results of these surveys, please indicate the manufacture, model number and sensitivity of the survey instrument used to measure residual radioactivity.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 103859.

Sincerely,

Original Signed By:
John E. Glenn

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety
and Safeguards

Enclosures:
Guidelines for Decontamination of Facilities

RI:DRSS *mg* RI:DRSS
Joustra/sm Glenn
6/19/85 6/17/85

8512200288 850724
REG1 LIC30
06-01060-01 PDR

OFFICIAL RECORD COPY

948JOUSTR6/6/85 - 0001.0.0
06/17/85

ML10